

LB 99428


245.1

INSTRUMENT ID: CX1

Linear

 μ Abs.:

35751



A= 0.0000e+000

B= 2.7791e-004

C= 2.9188e-002

Rho= 0.9999479

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	-68	0.000	-68				
0.05	0.050	0.063	0.013	120	0.0 %	120				
0.2	0.200	0.190	-0.010	580	0.0 %	580				
2.5	2.500	2.435	-0.065	8657	0.0 %	8657				
5.0	5.000	5.059	0.059	18097	0.0 %	18097				
7.5	7.500	7.528	0.028	26984	0.0 %	26984				
10.0	10.000	9.965	-0.035	35751	0.0 %	35751				

LB99428 INSTRUMENT ID : CV1

Method: 245.1		Operator: Admin		Date of Analysis: 28 Nov 2018 08:29:26					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-68	-	0.0000	245.1	PPB	28 Nov 2018 08:35:00	S	Std
0.05 - 1	50.5	120	-	0.0500	245.1	PPB	28 Nov 2018 08:39:55	S	Std
0.2 - 1	50.2	580	-	0.2000	245.1	PPB	28 Nov 2018 08:42:13	S	Std
2.5 - 1	52.5	8657	-	2.5000	245.1	PPB	28 Nov 2018 08:44:30	S	Std
5.0 - 1	55	18097	-	5.0000	245.1	PPB	28 Nov 2018 08:46:48	S	Std
7.5 - 1	57.5	26984	-	7.5000	245.1	PPB	28 Nov 2018 08:49:05	S	Std
10.0 - 1	510	35751	-	10.0000	245.1	PPB	28 Nov 2018 08:56:38	S	Std
ICV92 - 1	ICV92	14010	3.9227	-245.1		PPB	28 Nov 2018 09:00:17	U	SMPL
ICB92 - 1	ICB92	-213	-0.0300	-245.1		PPB	28 Nov 2018 09:02:33	U	SMPL
CCV05 - 1	CCV05	18717	5.2309	-245.1		PPB	28 Nov 2018 09:04:49	U	SMPL
CCB05 - 1	CCB05	-234	-0.0358	-245.1		PPB	28 Nov 2018 09:07:05	U	SMPL
CRA - 1	CRA	84	0.0525	-245.1		PPB	28 Nov 2018 09:09:22	U	SMPL
HighStd - 1	HighStd	35494	9.8934	-245.1		PPB	28 Nov 2018 09:11:37	U	SMPL
ChkStd - 1	ChkStd	24420	6.8158	-245.1		PPB	28 Nov 2018 09:13:54	U	SMPL
PB115111BL - 1	PBW	-216	-0.0308	-245.1		PPB	28 Nov 2018 09:16:13	U	SMPL
PB115111BS - 1	LCSW	14272	3.9955	-245.1		PPB	28 Nov 2018 09:18:31	U	SMPL
J6067-12 - 1	14B-(1-4)-COMP	844	0.2637	-245.1		PPB	28 Nov 2018 09:20:48	U	SMPL
J6069-01 - 1	112318-DBRI-F-D-B	-152	-0.0131	-245.1		PPB	28 Nov 2018 09:23:05	U	SMPL
J6069-02 - 1	112318-DBRI-F-D-M	-94	0.0031	-245.1		PPB	28 Nov 2018 09:25:23	U	SMPL
J6069-03 - 1	112318-DBRI-F-D-S	-94	0.0031	-245.1		PPB	28 Nov 2018 09:27:41	U	SMPL
J6069-04 - 1	112318-DBRI-F-U-B	-59	0.0128	-245.1		PPB	28 Nov 2018 09:29:56	U	SMPL
CCV06 - 1	CCV06	18163	5.0769	-245.1		PPB	28 Nov 2018 09:32:12	U	SMPL
CCB06 - 1	CCB06	-197	-0.0256	-245.1		PPB	28 Nov 2018 09:34:28	U	SMPL
J6069-05 - 1	112318-DBRI-F-U-M	-105	0.0000	-245.1		PPB	28 Nov 2018 09:36:45	U	SMPL
J6069-06 - 1	112318-DBRI-F-U-S	-90	0.0042	-245.1		PPB	28 Nov 2018 09:39:01	U	SMPL
J6069-07 - 1	112318-TIPC-F-D-B	-85	0.0056	-245.1		PPB	28 Nov 2018 09:41:17	U	SMPL
J6069-08 - 1	112318-TIPC-F-D-S	-92	0.0036	-245.1		PPB	28 Nov 2018 09:43:33	U	SMPL
J6069-09 - 1	112318-TIPC-F-U-B	-96	0.0025	-245.1		PPB	28 Nov 2018 09:45:50	U	SMPL
J6069-10 - 1	112318-TIPC-F-U-S	-71	0.0095	-245.1		PPB	28 Nov 2018 09:48:07	U	SMPL
J6069-11 - 1	112318-CHDE-E-D-B	-92	0.0036	-245.1		PPB	28 Nov 2018 09:50:25	U	SMPL
J6069-12 - 1	112318-CHDE-E-D-M	-98	0.0020	-245.1		PPB	28 Nov 2018 09:52:42	U	SMPL
J6069-13 - 1	112318-CHDE-E-D-S	-85	0.0056	-245.1		PPB	28 Nov 2018 09:55:00	U	SMPL
J6069-14 - 1	112318-CHDE-E-U-B	-108	-0.0008	-245.1		PPB	28 Nov 2018 09:57:16	U	SMPL
CCV07 - 1	CCV07	18325	5.1219	-245.1		PPB	28 Nov 2018 09:59:33	U	SMPL
CCB07 - 1	CCB07	-194	-0.0247	-245.1		PPB	28 Nov 2018 10:01:49	U	SMPL
J6069-15 - 1	112318-CHDE-E-U-M	-109	-0.0011	-245.1		PPB	28 Nov 2018 10:04:07	U	SMPL
J6069-16 - 1	112318-CHDE-E-U-S	-97	0.0022	-245.1		PPB	28 Nov 2018 10:06:23	U	SMPL
J6069-16DUP - 1	112318-CHDE-E-U-SDUP	-95	0.0028	-245.1		PPB	28 Nov 2018 10:08:39	U	SMPL
J6069-16MS - 1	112318-CHDE-E-U-SMS	16195	4.5300	-245.1		PPB	28 Nov 2018 10:10:56	U	SMPL
J6069-16MSD - 1	112318-CHDE-E-U-SMSD	16210	4.5341	-245.1		PPB	28 Nov 2018 10:13:13	U	SMPL
PB115112BL - 1	PBW	-79	0.0072	-245.1		PPB	28 Nov 2018 10:15:29	U	SMPL
PB115112BS - 1	LCSW	14100	3.9477	-245.1		PPB	28 Nov 2018 10:17:47	U	SMPL
J6070-01 - 1	112118-TIPC-E-D-M	-166	-0.0169	-245.1		PPB	28 Nov 2018 10:20:04	U	SMPL
J6070-02 - 1	112118-TIPC-E-U-M	-126	-0.0058	-245.1		PPB	28 Nov 2018 10:22:22	U	SMPL
J6070-03 - 1	112118-DBRI-E-D-B	-103	0.0006	-245.1		PPB	28 Nov 2018 10:24:39	U	SMPL
CCV08 - 1	CCV08	18353	5.1297	-245.1		PPB	28 Nov 2018 10:26:55	U	SMPL
CCB08 - 1	CCB08	-175	-0.0194	-245.1		PPB	28 Nov 2018 10:29:12	U	SMPL
J6070-04 - 1	112118-DBRI-E-D-M	-140	-0.0097	-245.1		PPB	28 Nov 2018 10:31:29	U	SMPL
J6070-05 - 1	112118-DBRI-E-D-S	-91	0.0039	-245.1		PPB	28 Nov 2018 10:33:46	U	SMPL
J6070-06 - 1	112118-DBRI-E-U-B	-83	0.0061	-245.1		PPB	28 Nov 2018 10:36:02	U	SMPL
J6070-07 - 1	112118-DBRI-E-U-M	-85	0.0056	-245.1		PPB	28 Nov 2018 10:38:19	U	SMPL
J6070-08 - 1	112118-DBRI-E-U-S	-102	0.0008	-245.1		PPB	28 Nov 2018 10:40:36	U	SMPL
J6070-09 - 1	112118-CHDE-E-D-B	-106	-0.0003	-245.1		PPB	28 Nov 2018 10:42:53	U	SMPL
J6070-10 - 1	112118-CHDE-E-D-M	-104	0.0003	-245.1		PPB	28 Nov 2018 10:45:11	U	SMPL
J6070-11 - 1	112118-CHDE-E-D-S	-102	0.0008	-245.1		PPB	28 Nov 2018 10:47:29	U	SMPL
J6070-12 - 1	112118-CHDE-E-U-B	-89	0.0045	-245.1		PPB	28 Nov 2018 10:49:47	U	SMPL
J6070-13 - 1	112118-CHDE-E-U-M	-87	0.0050	-245.1		PPB	28 Nov 2018 10:52:03	U	SMPL
CCV09 - 1	CCV09	18531	5.1792	-245.1		PPB	28 Nov 2018 10:54:20	U	SMPL
CCB09 - 1	CCB09	-185	-0.0222	-245.1		PPB	28 Nov 2018 10:56:36	U	SMPL
J6070-14 - 1	112118-CHDE-E-U-S	-112	-0.0019	-245.1		PPB	28 Nov 2018 10:58:53	U	SMPL
J6071-01 - 1	112418-CHDE-F-D-B	-99	0.0017	-245.1		PPB	28 Nov 2018 11:01:10	U	SMPL
J6071-02 - 1	112418-CHDE-F-D-M	-88	0.0047	-245.1		PPB	28 Nov 2018 11:03:26	U	SMPL
J6071-03 - 1	112418-CHDE-F-D-S	-98	0.0020	-245.1		PPB	28 Nov 2018 11:05:43	U	SMPL
J6071-04 - 1	112418-CHDE-F-U-B	-90	0.0042	-245.1		PPB	28 Nov 2018 11:08:00	U	SMPL
J6071-05 - 1	112418-CHDE-F-U-M	-86	0.0053	-245.1		PPB	28 Nov 2018 11:10:17	U	SMPL
J6071-06 - 1	112418-CHDE-F-U-S	-30	0.0209	-245.1		PPB	28 Nov 2018 11:12:34	U	SMPL
J6071-06DUP - 1	112418-CHDE-F-U-SDUP	-36	0.0192	-245.1		PPB	28 Nov 2018 11:14:52	U	SMPL
J6071-06MS - 1	112418-CHDE-F-U-SMS	15640	4.3757	-245.1		PPB	28 Nov 2018 11:17:10	U	SMPL
J6071-06MSD - 1	112418-CHDE-F-U-SMSD	15686	4.3885	-245.1		PPB	28 Nov 2018 11:19:26	U	SMPL
CCV10 - 1	CCV10	18567	5.1892	-245.1		PPB	28 Nov 2018 11:21:43	U	SMPL
CCB10 - 1	CCB10	-230	-0.0347	-245.1		PPB	28 Nov 2018 11:24:00	U	SMPL
PB115113BL - 1	PBW	-12	0.0259	-245.1		PPB	28 Nov 2018 11:26:17	U	SMPL
PB115113BS - 1	LCSW	15028	4.2056	-245.1		PPB	28 Nov 2018 11:28:34	U	SMPL
J6078-01 - 1	112618-JETT-F-D-B	-162	-0.0158	-245.1		PPB	28 Nov 2018 11:30:51	U	SMPL
J6078-02 - 1	112618-JETT-F-D-M	-52	0.0147	-245.1		PPB	28 Nov 2018 11:33:07	U	SMPL
J6078-03 - 1	112618-JETT-F-D-S	-82	0.0064	-245.1		PPB	28 Nov 2018 11:35:24	U	SMPL
J6078-04 - 1	112618-JETT-F-U-B	-75	0.0083	-245.1		PPB	28 Nov 2018 11:37:41	U	SMPL
J6078-05 - 1	112618-JETT-F-U-M	-33	0.0200	-245.1		PPB	28 Nov 2018 11:39:58	U	SMPL
J6078-06 - 1	112618-JETT-F-U-S	-37	0.0189	-245.1		PPB	28 Nov 2018 11:42:16	U	SMPL
J6078-07 - 1	112618-TIPC-F-D-B	-83	0.0061	-245.1		PPB	28 Nov 2018 11:44:34	U	SMPL
J6078-08 - 1	112618-TIPC-F-D-S	-83	0.0061	-245.1		PPB	28 Nov 2018 11:46:51	U	SMPL

LB99428 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Nov 2018 08:29:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCV11 - 1	CCV11	18301	5.1153		-245.1	PPB	28 Nov 2018 11:49:08	U	SMPL
CCB11 - 1	CCB11	-171	-0.0183		-245.1	PPB	28 Nov 2018 11:51:25	U	SMPL
J6078-09 - 1	112618-TIPC-F-U-B	-88	0.0047		-245.1	PPB	28 Nov 2018 11:53:43	U	SMPL
J6078-10 - 1	112618-TIPC-F-U-S	-79	0.0072		-245.1	PPB	28 Nov 2018 11:56:00	U	SMPL
J6078-11 - 1	112618-DBRI-F-D-B	-80	0.0070		-245.1	PPB	28 Nov 2018 11:58:17	U	SMPL
J6078-12 - 1	112618-DBRI-F-D-S	-92	0.0036		-245.1	PPB	28 Nov 2018 12:00:34	U	SMPL
J6078-13 - 1	112618-DBRI-F-U-B	-80	0.0070		-245.1	PPB	28 Nov 2018 12:02:51	U	SMPL
J6078-14 - 1	112618-DBRI-F-U-S	-66	0.0108		-245.1	PPB	28 Nov 2018 12:05:09	U	SMPL
J6078-15 - 1	112618-CHDE-F-D-B	-11	0.0261		-245.1	PPB	28 Nov 2018 12:07:26	U	SMPL
J6078-16 - 1	112618-CHDE-F-D-M	-69	0.0100		-245.1	PPB	28 Nov 2018 12:09:43	U	SMPL
J6078-17 - 1	112618-CHDE-F-D-S	25	0.0361		-245.1	PPB	28 Nov 2018 12:12:01	U	SMPL
J6078-18 - 1	112618-CHDE-F-U-B	-68	0.0103		-245.1	PPB	28 Nov 2018 12:14:18	U	SMPL
CCV12 - 1	CCV12	18296	5.1139		-245.1	PPB	28 Nov 2018 12:16:35	U	SMPL
CCB12 - 1	CCB12	-170	-0.0181		-245.1	PPB	28 Nov 2018 12:18:53	U	SMPL
J6078-19 - 1	112618-CHDE-F-U-M	-73	0.0089		-245.1	PPB	28 Nov 2018 12:21:10	U	SMPL
J6078-20 - 1	112618-CHDE-F-U-S	-72	0.0092		-245.1	PPB	28 Nov 2018 12:23:27	U	SMPL
J6078-20DUP - 1	112618-CHDE-F-U-SDUP	-62	0.0120		-245.1	PPB	28 Nov 2018 12:25:44	U	SMPL
J6078-20MS - 1	112618-CHDE-F-U-SMS	16065	4.4938		-245.1	PPB	28 Nov 2018 12:28:02	U	SMPL
J6078-20MSD - 1	112618-CHDE-F-U-SMSD	15930	4.4563		-245.1	PPB	28 Nov 2018 12:30:19	U	SMPL
CCV13 - 1	CCV13	17989	5.0285		-245.1	PPB	28 Nov 2018 12:32:36	U	SMPL
CCB13 - 1	CCB13	-193	-0.0244		-245.1	PPB	28 Nov 2018 12:34:53	U	SMPL
J6069-16LX5 - 1		-105	0.0000		-245.1	PPB	28 Nov 2018 12:37:11	U	SMPL
J6069-16A - 1		17247	4.8223		-245.1	PPB	28 Nov 2018 12:39:29	U	SMPL
J6071-06LX5 - 1		-134	-0.0081		-245.1	PPB	28 Nov 2018 12:41:47	U	SMPL
J6071-06A - 1		17457	4.8807		-245.1	PPB	28 Nov 2018 12:44:04	U	SMPL
J6078-20LX5 - 1		-158	-0.0147		-245.1	PPB	28 Nov 2018 12:46:21	U	SMPL
J6078-20A - 1		17068	4.7726		-245.1	PPB	28 Nov 2018 12:48:39	U	SMPL
CCV14 - 1	CCV14	18429	5.1508		-245.1	PPB	28 Nov 2018 12:50:56	U	SMPL
CCB14 - 1	CCB14	-181	-0.0211		-245.1	PPB	28 Nov 2018 12:53:13	U	SMPL