



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99430

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.5	4.01	11/27/2018	16:15
2	CAL2	1	Water			20.6	7.00	11/27/2018	16:17
3	CAL3	1	Water			20.4	10.03	11/27/2018	16:19
4	ICV	1	Water			20.4	7.01	11/27/2018	16:21
5	CCV1	1	Water			20.5	2.00	11/27/2018	16:23
6	J6073-01	1	Solid	20.02	20	21.2	7.40	11/27/2018	16:25
7	J6073-01DUP	1	Solid	20.03	20	21.3	7.41	11/27/2018	16:27
8	J6073-02	1	Solid	20.03	20	21.4	6.83	11/27/2018	16:29
9	J6083-02	1	Solid	20.02	20	20.7	7.87	11/27/2018	16:31
10	J6083-04	1	Solid	20.04	20	21.0	8.96	11/27/2018	16:33
11	J6083-06	1	Solid	20.03	20	21.0	9.07	11/27/2018	16:35
12	J6085-03	1	Solid	20.02	20	21.0	5.60	11/27/2018	16:37
13	CCV2	1	Water			20.4	12.00	11/27/2018	16:45

WORKLIST(Hardcopy Internal Chain)

LB 99430

WorkList Name : corrosivity112718

WorkList ID : 119627

Date : 11/27/2018 3:28:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/21/2018	Solid	J6073-01	Corrosivity	Cool 4 deg C	CLEA11	L13	CHASIS-WASH	11/21/2018	9045D
11/21/2018	Solid	J6073-02	Corrosivity	Cool 4 deg C	CLEA11	L13	OWS-2	11/21/2018	9045D
11/27/2018	Solid	J6083-02	Corrosivity	Cool 4 deg C	AREC01	L43	WC-1	11/27/2018	9045D
11/27/2018	Solid	J6083-04	Corrosivity	Cool 4 deg C	AREC01	L43	WC-2	11/27/2018	9045D
11/27/2018	Solid	J6083-06	Corrosivity	Cool 4 deg C	AREC01	L43	WC-3	11/27/2018	9045D
11/27/2018	Solid	J6085-03	Corrosivity	Cool 4 deg C	PSEG01	L31	NS-OILY-STONE	11/27/2018	9045D

Date/Time 11/27/18 4:00
 Received by: SP
 Relinquished by: CP

Date/Time 11/27/18 4:55
 Received by: CP
 Relinquished by: SP