



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99431

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.5	4.01	11/27/2018	16:15
2	CAL2	1	Water			20.6	7.00	11/27/2018	16:17
3	CAL3	1	Water			20.4	10.03	11/27/2018	16:19
4	ICV	1	Water			20.4	7.01	11/27/2018	16:21
5	CCV1	1	Water			20.5	2.00	11/27/2018	16:23
6	J6084-02	1	Water			12.0	10.64	11/27/2018	16:39
7	J6084-02DUP	1	Water			12.1	10.65	11/27/2018	16:41
8	J6085-05	1	Water			13.3	6.62	11/27/2018	16:43
9	CCV2	1	Water			20.5	12.00	11/27/2018	16:45

WORKLIST(Hardcopy Internal Chain)

JB 99431

WorkList Name : corrosivity112718-1

WorkList ID : 119628

Date : 11/27/2018 3:28:51 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/27/2018	Water	J6084-02	Corrosivity	Cool 4 deg C	PSEG01	L31	MS-SFVP8283LM	11/27/2018	9040C
11/27/2018	Water	J6085-05	Corrosivity	Cool 4 deg C	PSEG01	L31	NS	11/27/2018	9040C

Date/Time 11/27/18 4:00
 Received by: SP
 Relinquished by: CP

Date/Time 11/27/18 4:55
 Received by: SP
 Relinquished by: SP