



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB99440
Analysis Date: 11/28/2018
BalanceID: WC SC-3
OvenID: EX OVEN-2

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 11/28/2018
Extraction IN Time: 07:40
Extraction OUT Time: 08:35

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB99440BL	LB99440BL	WATER	2	1000	100	2.6147	2.6147	3.01	2.6147	2.6147	0.0000	0
2	LB99440BS	LB99440BS	WATER	2	1000	100	2.8634	2.8634	3.02	2.8819	2.8819	0.0185	18.5
3	LB99440BSD	LB99440BSD	WATER	2	1000	100	2.8103	2.8103	3.03	2.8290	2.8290	0.0187	18.7
4	J6064-03	OILY-WATER	WATER	2	1000	100	2.9824	2.9824	3.04	3.0218	3.0218	0.0394	39.4
5	J6084-01	MS-SFVP8283LM	WATER	2	1000	100	3.0814	3.0814	3.05	3.0990	3.0990	0.0176	17.6
6	J6085-04	NS	WATER	2	1000	100	3.2076	3.2076	18.04	3.4172	3.4172	0.2096	209.6



QC Batch# LB99440 Test: TPH Analysis Date: 11/28/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2448
pH Paper 0-14	W2416
Sodium Sulfate	EP1897
1:1 HCL	WP69312
Silica Gel	W2368
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP69314
LCSWD	10.00 ML	WP69313
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0018</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>10:16</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>09:15</u>		
Bal Check Time:	<u>08:15</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:40</u>
		Out Time:	<u>10:15</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:51</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>11:15</u>		
Bal Check Time:	<u>12:50</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>12:25</u>
		Out Time:	<u>11:50</u>		

1309440

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH J6084

WorkList ID : 119635

Date : 11/28/2018 7:30:51 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/30/2018	Water	J6064-03	-K- TPH	Conc H2SO4 tr	PSEG01	L31	OILY-WATER	11/26/2018	1664A
11/30/2018	Water	J6084-01	- < TPH	Conc H2SO4 tr	PSEG01	L31	MS-SFVP8283LM	11/27/2018	1664A
11/30/2018	Water	J6085-04	- < TPH	Conc H2SO4 tr	PSEG01	L31	NS	11/27/2018	1664A

Date/Time 11-28-18 7:35 AM
 Received by: JP
 Relinquished by: CJ

Date/Time 11-28-18 10:20 AM
 Received by: CJ
 Relinquished by: JP