


LB99446

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	41	0.000	41				
0.2	0.200	0.161	-0.039	658	0.0 %	658				
2.5	2.500	2.434	-0.066	9848	0.0 %	9848				
5.0	5.000	5.169	0.169	20902	0.0 %	20902				
7.5	7.500	7.496	-0.004	30311	0.0 %	30311				
10.0	10.000	9.932	-0.068	40156	0.0 %	40156				

LB99446 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 28 Nov 2018 14:27:23					
Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.0000	7470A	PPB	28 Nov 2018 14:29:15	S	Std
0.2 - 1	50.2	658	-	0.2000	7470A	PPB	28 Nov 2018 14:31:24	S	Std
2.5 - 1	52.5	9848	-	2.5000	7470A	PPB	28 Nov 2018 14:33:32	S	Std
5.0 - 1	55	20902	-	5.0000	7470A	PPB	28 Nov 2018 14:35:41	S	Std
7.5 - 1	57.5	30311	-	7.5000	7470A	PPB	28 Nov 2018 14:37:49	S	Std
10.0 - 1	510	40156	-	10.0000	7470A	PPB	28 Nov 2018 14:42:29	S	Std
ICV93 - 1	ICV93	15071	3.7263	-	7470A	PPB	28 Nov 2018 14:47:26	U	SMPL
ICB93 - 1	ICB93	-38	-0.0113	-	7470A	PPB	28 Nov 2018 14:49:33	U	SMPL
CCV15 - 1	CCV15	20349	5.0319	-	7470A	PPB	28 Nov 2018 14:54:15	U	SMPL
CCB15 - 1	CCB15	-60	-0.0168	-	7470A	PPB	28 Nov 2018 14:56:21	U	SMPL
CRA - 1	CRA	775	0.1898	-	7470A	PPB	28 Nov 2018 14:58:28	U	SMPL
HighStd - 1	HighStd	40454	10.0054	-	7470A	PPB	28 Nov 2018 15:00:35	U	SMPL
ChkStd - 1	ChkStd	28119	6.9541	-	7470A	PPB	28 Nov 2018 15:02:42	U	SMPL
PB115105BL - 1	PBW	-106	-0.0282	-	7470A	PPB	28 Nov 2018 15:04:50	U	SMPL
PB115105BS - 1	LCSW	14746	3.6459	-	7470A	PPB	28 Nov 2018 15:06:59	U	SMPL
J6062-03 - 1	B-20(1-10)	0	-0.0019	-	7470A	PPB	28 Nov 2018 15:09:07	U	SMPL
J6062-03DUP - 1	B-20(1-10)DUP	18	0.0025	-	7470A	PPB	28 Nov 2018 15:11:16	U	SMPL
J6062-03MS - 1	B-20(1-10)MS	13489	3.3349	-	7470A	PPB	28 Nov 2018 15:13:24	U	SMPL
J6062-03MSD - 1	B-20(1-10)MSD	13604	3.3634	-	7470A	PPB	28 Nov 2018 15:15:33	U	SMPL
J6064-02 - 1	OILY-SOIL	-63	-0.0175	-	7470A	PPB	28 Nov 2018 15:17:40	U	SMPL
CCV16 - 1	CCV16	19123	4.7286	-	7470A	PPB	28 Nov 2018 15:19:46	U	SMPL
CCB16 - 1	CCB16	-69	-0.0190	-	7470A	PPB	28 Nov 2018 15:21:53	U	SMPL
J6065-02 - 1	USTS-1-2-3	20	0.0030	-	7470A	PPB	28 Nov 2018 15:23:59	U	SMPL
PB115106BL - 1	PBW	25	0.0042	-	7470A	PPB	28 Nov 2018 15:26:06	U	SMPL
PB115106BS - 1	LCSW	14131	3.4937	-	7470A	PPB	28 Nov 2018 15:28:13	U	SMPL
J6064-04 - 1	OILY-WATER	6	-0.0005	-	7470A	PPB	28 Nov 2018 15:30:20	U	SMPL
J6064-04DUP - 1	OILY-WATERDUP	29	0.0052	-	7470A	PPB	28 Nov 2018 15:32:27	U	SMPL
J6064-04MS - 1	OILY-WATERMS	14956	3.6978	-	7470A	PPB	28 Nov 2018 15:37:21	U	SMPL
J6064-04MSD - 1	OILY-WATERMSD	14941	3.6941	-	7470A	PPB	28 Nov 2018 15:39:29	U	SMPL
PB115107BL - 1	PBW	-8	-0.0039	-	7470A	PPB	28 Nov 2018 15:41:37	U	SMPL
PB115107BS - 1	LCSW	13290	3.2857	-	7470A	PPB	28 Nov 2018 15:43:46	U	SMPL
J6060-08 - 1	SITE-WATER	-132	-0.0346	-	7470A	PPB	28 Nov 2018 15:45:53	U	SMPL
CCV17 - 1	CCV17	21651	5.3540	-	7470A	PPB	28 Nov 2018 15:48:00	U	SMPL
CCB17 - 1	CCB17	-59	-0.0165	-	7470A	PPB	28 Nov 2018 15:50:07	U	SMPL
J6060-08DUP - 1	SITE-WATERDUP	-92	-0.0247	-	7470A	PPB	28 Nov 2018 15:52:14	U	SMPL
J6060-08MS - 1	SITE-WATERMS	13382	3.3085	-	7470A	PPB	28 Nov 2018 15:54:22	U	SMPL
J6060-08MSD - 1	SITE-WATERMSD	14253	3.5239	-	7470A	PPB	28 Nov 2018 15:58:58	U	SMPL
J6064-03 - 1	OILY-WATER	3	-0.0012	-	7470A	PPB	28 Nov 2018 16:01:05	U	SMPL
PB115106TB - 1	PB115075TB	-85	-0.0230	-	7470A	PPB	28 Nov 2018 16:05:19	U	SMPL
PB115105TB - 1	PB115073TB	16	0.0020	-	7470A	PPB	28 Nov 2018 16:07:56	U	SMPL
CCV18 - 1	CCV18	20326	5.0262	-	7470A	PPB	28 Nov 2018 16:10:22	U	SMPL
CCB18 - 1	CCB18	-51	-0.0146	-	7470A	PPB	28 Nov 2018 16:12:30	U	SMPL
J6062-03LX5 - 1		9	0.0003	-	7470A	PPB	28 Nov 2018 16:14:39	U	SMPL
J6062-03A - 1		14669	3.6268	-	7470A	PPB	28 Nov 2018 16:16:46	U	SMPL
J6064-04LX5 - 1		-54	-0.0153	-	7470A	PPB	28 Nov 2018 16:18:53	U	SMPL
J6064-04A - 1		14053	3.4744	-	7470A	PPB	28 Nov 2018 16:21:00	U	SMPL
J6060-08LX5 - 1		-69	-0.0190	-	7470A	PPB	28 Nov 2018 16:23:08	U	SMPL
J6060-08A - 1		14828	3.6662	-	7470A	PPB	28 Nov 2018 16:25:15	U	SMPL
CCV19 - 1	CCV19	20129	4.9775	-	7470A	PPB	28 Nov 2018 16:27:22	U	SMPL
CCB19 - 1	CCB19	-52	-0.0148	-	7470A	PPB	28 Nov 2018 16:29:29	U	SMPL