



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99453

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 051106

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 7                 | 27.0       | 80.6      | 81.2            | 11/28/2018 | 10:05     |
| 2   | J6084-02    |               | 1  | 9                 | >100.0     | >212.0    | >212.0          | 11/28/2018 | 10:35     |
| 3   | J6084-02DUP |               | 1  | 9                 | >100.0     | >212.0    | >212.0          | 11/28/2018 | 10:55     |
| 4   | J6085-05    |               | 1  | 8                 | >100.0     | >212.0    | >212.0          | 11/28/2018 | 11:20     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

LB99453

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp112818

WorkList ID : 119659

Date : 11/28/2018 9:20:20 AM

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
| 11/30/2018 | Water  | J6084-02 | Flash Point | Cool 4 deg C | PSEG01   | L31              | MS-SFVP8283LM   | 11/27/2018   | 1010A  |
| 11/30/2018 | Water  | J6085-05 | Flash Point | Cool 4 deg C | PSEG01   | L31              | NS              | 11/27/2018   | 1010A  |

Date/Time 11/28/18 9:30 AM

Received by: KS

Relinquished by: CP

Date/Time 11/28/18 11:30 AM

Received by: CP

Relinquished by: KS