


LB99454

7471B

INSTRUMENT ID: CV1

 μ Abs.:

52840



Linear

A= 0.0000e+000

B= 1.8878e-004

C= 3.3852e-002

Rho= 0.9998016

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.047	0.047	70	0.000	70				
0.2	0.200	0.206	0.006	912	0.0 %	912				
2.5	2.500	2.376	-0.124	12407	0.0 %	12407				
5.0	5.000	5.111	0.111	26895	0.0 %	26895				
7.5	7.500	7.451	-0.049	39291	0.0 %	39291				
10.0	10.000	10.009	0.009	52840	0.0 %	52840				

LB99454 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Nov 2018 16:50:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	70	-	0.0000	7471B	PPB	28 Nov 2018 16:52:27	S	Std
0.2 - 1	50.2	912	-	0.2000	7471B	PPB	28 Nov 2018 16:54:45	S	Std
2.5 - 1	52.5	12407	-	2.5000	7471B	PPB	28 Nov 2018 16:57:03	S	Std
5.0 - 1	55	26895	-	5.0000	7471B	PPB	28 Nov 2018 16:59:21	S	Std
7.5 - 1	57.5	39291	-	7.5000	7471B	PPB	28 Nov 2018 17:04:43	S	Std
10.0 - 1	59.0	52840	-	10.0000	7471B	PPB	28 Nov 2018 17:09:36	S	Std
ICV94 - 1	ICV94	21237	4.0429	-	7471B	PPB	28 Nov 2018 17:12:21	U	SMPL
ICB94 - 1	ICB94	-83	0.0182	-	7471B	PPB	28 Nov 2018 17:14:37	U	SMPL
CCV20 - 1	CCV20	26743	5.0823	-	7471B	PPB	28 Nov 2018 17:16:54	U	SMPL
CCB20 - 1	CCB20	55	0.0442	-	7471B	PPB	28 Nov 2018 17:21:43	U	SMPL
CRA - 1	CRA	880	0.2000	-	7471B	PPB	28 Nov 2018 17:23:58	U	SMPL
HighStd - 1	HighStd	53360	10.1070	-	7471B	PPB	28 Nov 2018 17:26:14	U	SMPL
ChkStd - 1	ChkStd	37691	7.1490	-	7471B	PPB	28 Nov 2018 17:28:30	U	SMPL
PB115115BL - 1	PBS	-64	0.0218	-	7471B	PPB	28 Nov 2018 17:30:48	U	SMPL
PB115115BS - 1	LCSS	18966	3.6142	-	7471B	PPB	28 Nov 2018 17:33:06	U	SMPL
J6074-01 - 1	001-WC-NY5	1858	0.3846	-	7471B	PPB	28 Nov 2018 17:35:23	U	SMPL
J6074-01DUP - 1	001-WC-NY5DUP	1865	0.3859	-	7471B	PPB	28 Nov 2018 17:37:41	U	SMPL
J6074-01MS - 1	001-WC-NY5MS	21233	4.0422	-	7471B	PPB	28 Nov 2018 17:39:59	U	SMPL
J6074-01MSD - 1	001-WC-NY5MSD	21157	4.0278	-	7471B	PPB	28 Nov 2018 17:42:17	U	SMPL
J6074-03 - 1	002-WC-NY6	1936	0.3993	-	7471B	PPB	28 Nov 2018 17:44:33	U	SMPL
CCV21 - 1	CCV21	26748	5.0833	-	7471B	PPB	28 Nov 2018 17:46:49	U	SMPL
CCB21 - 1	CCB21	-75	0.0197	-	7471B	PPB	28 Nov 2018 17:49:05	U	SMPL
J6074-05 - 1	003-WC-NY2	5141	1.0044	-	7471B	PPB	28 Nov 2018 17:51:22	U	SMPL
J6074-07 - 1	004-WC-NY4	770	0.1792	-	7471B	PPB	28 Nov 2018 17:53:38	U	SMPL
J6079-01 - 1	J-1	3248	0.6470	-	7471B	PPB	28 Nov 2018 17:55:54	U	SMPL
J6079-02 - 1	J-2	1236	0.2672	-	7471B	PPB	28 Nov 2018 17:58:10	U	SMPL
J6079-03 - 1	J-3	5574	1.0861	-	7471B	PPB	28 Nov 2018 18:00:27	U	SMPL
J6079-04 - 1	J-4	6581	1.2762	-	7471B	PPB	28 Nov 2018 18:02:44	U	SMPL
J6080-01 - 1	RCA-04	7508	1.4512	-	7471B	PPB	28 Nov 2018 18:05:01	U	SMPL
J6080-02 - 1	RCA-05	9662	1.8578	-	7471B	PPB	28 Nov 2018 18:07:18	U	SMPL
J6080-03 - 1	RCA-06	7818	1.5097	-	7471B	PPB	28 Nov 2018 18:09:36	U	SMPL
J6080-04 - 1	RCA-07	10146	1.9492	-	7471B	PPB	28 Nov 2018 18:11:53	U	SMPL
CCV22 - 1	CCV22	27072	5.1444	-	7471B	PPB	28 Nov 2018 18:14:09	U	SMPL
CCB22 - 1	CCB22	-70	0.0206	-	7471B	PPB	28 Nov 2018 18:16:26	U	SMPL
J6083-01 - 1	WC-1	1344	0.2876	-	7471B	PPB	28 Nov 2018 18:18:43	U	SMPL
J6083-03 - 1	WC-2	2284	0.4650	-	7471B	PPB	28 Nov 2018 18:20:59	U	SMPL
J6083-05 - 1	WC-3	2891	0.5796	-	7471B	PPB	28 Nov 2018 18:23:16	U	SMPL
J6085-02 - 1	NS-OILY-STONE	222	0.0758	-	7471B	PPB	28 Nov 2018 18:25:33	U	SMPL
J6074-01LX5 - 1		395	0.1084	-	7471B	PPB	28 Nov 2018 18:27:50	U	SMPL
J6074-01A - 1		23004	4.3765	-	7471B	PPB	28 Nov 2018 18:30:07	U	SMPL
CCV23 - 1	CCV23	26630	5.0610	-	7471B	PPB	28 Nov 2018 18:32:24	U	SMPL
CCB23 - 1	CCB23	-57	0.0231	-	7471B	PPB	28 Nov 2018 18:34:42	U	SMPL