



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99458

Slope : 99.3

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.5	4.01	11/28/2018	12:30
2	CAL2	1	Water			20.6	7.00	11/28/2018	12:32
3	CAL3	1	Water			20.4	10.04	11/28/2018	12:34
4	ICV	1	Water			20.5	7.01	11/28/2018	12:36
5	CCV1	1	Water			20.6	2.00	11/28/2018	12:38
6	J6097-01	1	Water			15.3	5.43	11/28/2018	12:40
7	J6097-01DUP	1	Water			15.4	5.44	11/28/2018	12:42
8	CCV2	1	Water			20.4	12.00	11/28/2018	12:44

CB99458

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH112818

WorkList ID : 119669

Date : 11/28/2018 12:11:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/26/2018	Water	J6097-01	pH	Cool 4 deg C	STER12	L12	CELL-POOL-WATER-4TH-Q	11/26/2018	SM 4500-H B

Date/Time 11/28/18 12:15
Received by: SP
Relinquished by: CP

Date/Time 11/28/18 12:50
Received by: CP
Relinquished by: SP