



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99460

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	11/28/2018	17:30
2	CAL2	1	Water			21.3	7.00	11/28/2018	17:32
3	CAL3	1	Water			21.4	10.03	11/28/2018	17:34
4	ICV	1	Water			21.3	7.01	11/28/2018	17:36
5	CCV1	1	Water			21.4	2.00	11/28/2018	17:54
6	J6124-01	1	Water			20.7	7.62	11/28/2018	17:56
7	J6124-01DUP	1	Water			20.8	7.63	11/28/2018	17:58
8	CCV2	1	Water			21.3	12.00	11/28/2018	18:00

WORKLIST(Hardcopy Internal Chain)

LB 99460

WorkList Name : pH112818-1

WorkList ID : 119677

Date : 11/28/2018 4:36:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/28/2018	Water	J6124-01	pH	Cool 4 deg C	PSEG01	L31	BW-2072-112818	11/28/2018	9040C

Date/Time 11/28/18 5:25
 Received by: SP
 Relinquished by: CP

Date/Time 11/28/18 6:05
 Received by: CP
 Relinquished by: SP