


7470A

LB 99462
INSTRUMENT ID: CV2

Linear

 μ Abs.:

3755E



A= 0.0000e+000

B= 2.6125e-004

C= 1.1821e-002

Rho= 0.9994315

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.023	0.023	43	0.000	43				
0.2	0.200	0.178	-0.022	635	0.0 %	635				
2.5	2.500	2.396	-0.104	9127	0.0 %	9127				
5.0	5.000	5.064	0.064	19340	0.0 %	19340				
7.5	7.500	7.714	0.214	29483	0.0 %	29483				
10.0	10.000	9.824	-0.176	37558	0.0 %	37558				

LB99462 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Nov 2018 10:10:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	43	-	0.0000	7470A	PPB	29 Nov 2018 10:27:59	S	Std
0.2 - 1	50.2	635	-	0.2000	7470A	PPB	29 Nov 2018 10:30:07	S	Std
2.5 - 1	52.5	9127	-	2.5000	7470A	PPB	29 Nov 2018 10:32:16	S	Std
5.0 - 1	55	19340	-	5.0000	7470A	PPB	29 Nov 2018 10:34:24	S	Std
7.5 - 1	57.5	29483	-	7.5000	7470A	PPB	29 Nov 2018 10:36:32	S	Std
10.0 - 1	516	37558	-	10.0000	7470A	PPB	29 Nov 2018 10:41:07	S	Std
ICV95 - 1	ICV95	14891	3.9022	-	7470A	PPB	29 Nov 2018 10:43:47	U	SMPL
ICB95 - 1	ICB95	-57	-0.0031	-	7470A	PPB	29 Nov 2018 10:45:53	U	SMPL
CCV24 - 1	CCV24	19128	5.0091	-	7470A	PPB	29 Nov 2018 10:48:00	U	SMPL
CCB24 - 1	CCB24	-47	-0.0005	-	7470A	PPB	29 Nov 2018 10:50:07	U	SMPL
CRA - 1	CRA	602	0.1691	-	7470A	PPB	29 Nov 2018 10:52:13	U	SMPL
HighStd - 1	HighStd	37989	9.9366	-	7470A	PPB	29 Nov 2018 10:54:20	U	SMPL
ChkStd - 1	ChkStd	26591	6.9589	-	7470A	PPB	29 Nov 2018 10:56:27	U	SMPL
PB115145BL - 1	PBW	-77	-0.0083	-	7470A	PPB	29 Nov 2018 10:58:35	U	SMPL
PB115145BS - 1	LCSW	14872	3.8972	-	7470A	PPB	29 Nov 2018 11:00:43	U	SMPL
J6073-01 - 1	CHASIS-WASH	-31	0.0037	-	7470A	PPB	29 Nov 2018 11:02:51	U	SMPL
J6073-02 - 1	OWS-2	35	0.0210	-	7470A	PPB	29 Nov 2018 11:04:59	U	SMPL
J6074-02 - 1	001-WC-NY5	-3	0.0110	-	7470A	PPB	29 Nov 2018 11:07:07	U	SMPL
J6074-02DUP - 1	001-WC-NY5DUP	11	0.0147	-	7470A	PPB	29 Nov 2018 11:09:16	U	SMPL
J6074-02MS - 1	001-WC-NY5MS	14610	3.8288	-	7470A	PPB	29 Nov 2018 11:11:22	U	SMPL
CCV25 - 1	CCV25	19145	5.0135	-	7470A	PPB	29 Nov 2018 11:13:29	U	SMPL
CCB25 - 1	CCB25	54	0.0259	-	7470A	PPB	29 Nov 2018 11:15:35	U	SMPL
J6074-02MSD - 1	001-WC-NY5MSD	13410	3.5153	-	7470A	PPB	29 Nov 2018 11:17:42	U	SMPL
J6074-04 - 1	002-WC-NY6	-74	-0.0075	-	7470A	PPB	29 Nov 2018 11:19:49	U	SMPL
J6074-06 - 1	003-WC-NY2	-6	0.0103	-	7470A	PPB	29 Nov 2018 11:21:56	U	SMPL
J6074-08 - 1	004-WC-NY4	-9	0.0095	-	7470A	PPB	29 Nov 2018 11:24:04	U	SMPL
J6083-02 - 1	WC-1	39	0.0220	-	7470A	PPB	29 Nov 2018 11:26:11	U	SMPL
J6083-04 - 1	WC-2	21	0.0173	-	7470A	PPB	29 Nov 2018 11:28:19	U	SMPL
J6083-06 - 1	WC-3	17	0.0163	-	7470A	PPB	29 Nov 2018 11:30:27	U	SMPL
J6085-03 - 1	NS-OILY-STONE	31	0.0199	-	7470A	PPB	29 Nov 2018 11:32:35	U	SMPL
PB115146BL - 1	PBW	23	0.0178	-	7470A	PPB	29 Nov 2018 11:34:44	U	SMPL
PB115146BS - 1	LCSW	14240	3.7321	-	7470A	PPB	29 Nov 2018 11:36:51	U	SMPL
CCV26 - 1	CCV26	19027	4.9827	-	7470A	PPB	29 Nov 2018 11:38:58	U	SMPL
CCB26 - 1	CCB26	-48	-0.0007	-	7470A	PPB	29 Nov 2018 11:41:06	U	SMPL
J6084-02 - 1	MS-SFVP8283LM	7017	1.8450	-	7470A	PPB	29 Nov 2018 11:43:12	U	SMPL
J6085-05 - 1	NS	125	0.0445	-	7470A	PPB	29 Nov 2018 11:51:44	U	SMPL
J6085-05DUP - 1	NSDUP	144	0.0494	-	7470A	PPB	29 Nov 2018 11:53:51	U	SMPL
J6085-05MS - 1	NSMS	13814	3.6208	-	7470A	PPB	29 Nov 2018 11:55:58	U	SMPL
J6085-05MSD - 1	NSMSD	13484	3.5346	-	7470A	PPB	29 Nov 2018 11:58:06	U	SMPL
PB115147BL - 1	PBW	-34	0.0029	-	7470A	PPB	29 Nov 2018 12:00:13	U	SMPL
PB115147BS - 1	LCSW	13659	3.5803	-	7470A	PPB	29 Nov 2018 12:02:22	U	SMPL
J6084-01 - 1	MS-SFVP8283LM	99612	26.0360	-	7470A	PPB	29 Nov 2018 12:04:31	U	SMPL
J6085-04 - 1	NS	-31	0.0037	-	7470A	PPB	29 Nov 2018 12:09:18	U	SMPL
J6085-04DUP - 1	NSDUP	-58	-0.0033	-	7470A	PPB	29 Nov 2018 12:11:25	U	SMPL
CCV27 - 1	CCV27	19260	5.0436	-	7470A	PPB	29 Nov 2018 12:13:32	U	SMPL
CCB27 - 1	CCB27	-53	-0.0020	-	7470A	PPB	29 Nov 2018 12:15:39	U	SMPL
J6085-04MS - 1	NSMS	14613	3.8295	-	7470A	PPB	29 Nov 2018 12:20:16	U	SMPL
J6085-04MSD - 1	NSMSD	14839	3.8886	-	7470A	PPB	29 Nov 2018 12:22:22	U	SMPL
PB115145TB - 1	PB115108TB	-13	0.0084	-	7470A	PPB	29 Nov 2018 12:24:30	U	SMPL
PB115146TB - 1	PB115110TB	27	0.0189	-	7470A	PPB	29 Nov 2018 12:26:37	U	SMPL
J6074-02LX5 - 1		21	0.0173	-	7470A	PPB	29 Nov 2018 12:28:44	U	SMPL
J6074-02A - 1		14123	3.7015	-	7470A	PPB	29 Nov 2018 12:30:52	U	SMPL
J6085-05LX5 - 1		5	0.0131	-	7470A	PPB	29 Nov 2018 12:33:00	U	SMPL
J6085-05A - 1		14958	3.9197	-	7470A	PPB	29 Nov 2018 12:35:09	U	SMPL
J6085-04LX5 - 1		-50	-0.0012	-	7470A	PPB	29 Nov 2018 12:37:17	U	SMPL
J6085-04A - 1		12578	3.2979	-	7470A	PPB	29 Nov 2018 12:39:24	U	SMPL
CCV28 - 1	CCV28	19017	4.9801	-	7470A	PPB	29 Nov 2018 12:41:31	U	SMPL
CCB28 - 1	CCB28	-38	0.0019	-	7470A	PPB	29 Nov 2018 12:43:39	U	SMPL
J6084-01DLX5 - 1		19108	5.0039	-	7470A	PPB	29 Nov 2018 12:45:46	U	SMPL
CCV29 - 1	CCV29	19140	5.0122	-	7470A	PPB	29 Nov 2018 12:47:53	U	SMPL
CCB29 - 1	CCB29	-52	-0.0018	-	7470A	PPB	29 Nov 2018 12:50:00	U	SMPL