



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99469

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	11/29/2018	10:05
2	J6044-01		1	8	>100.0	>212.0	>212.0	11/29/2018	10:40
3	J6044-01DUP		1	8	>100.0	>212.0	>212.0	11/29/2018	11:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 99469

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp112918

WorkList ID : 119709

Date : 11/29/2018 9:27:25 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/28/2018	Solid	J6044-01	A Flash Point	Cool 4 deg C	PSEG01	L31	30055	11/21/2018	1010A

Date/Time 11/29/18 9:30 AM

Received by: KS

Relinquished by: CP

Date/Time

Received by:

Relinquished by:

11/29/18 11:10 AM

CP

KS