



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99470

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.9	11/29/2018	11:45
2	J6124-01		1	9	>100.0	>212.0	>212.0	11/29/2018	12:20
3	J6124-01DUP		1	9	>100.0	>212.0	>212.0	11/29/2018	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB99470

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp112918-1

WorkList ID : 119710

Date : 11/29/2018 9:29:57 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/03/2018	Water	J6124-01	F	Cool 4 deg C	PSEG01	L31	BW-2072-112818	11/28/2018	1010A

Date/Time 11/29/18 9:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 11/29/18 12:50 PM
 Received by: CP
 Relinquished by: KS