



Analytical Summary Report

Analysis Method: SM3500-Cr B
Parameter: Hexavalent Chromium
Run Number: LB99477

ANALYST: Indiana
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include Calibration Std. hexchrome 0.1 ppm, Hexavalent Chromium LCS Std, Hexavalent Chromium ICV Std, etc.

Intercept: -0 Slope: 0.7574 Regression: 0.999973

Table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.



Analytical Summary Report

Analysis Method: SM3500-Cr B

ANALYST: Indiana

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: apatel

Run Number: LB99477

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		2.28	0.000	0.375	0.375	0.495	11/29/2018	11:03
2	ICB		1	100	100		2.26	0.000	0.000	0.000	0.000	11/29/2018	11:04
3	CCV1	0.5	1	100	100		2.02	0.000	0.377	0.377	0.498	11/29/2018	17:30
4	CCB1		1	100	100		2.31	0.000	0.000	0.000	0.000	11/29/2018	17:30
5	LB99477BL		1	100	100		1.66	0.000	0.000	0.000	0.000	11/29/2018	17:31
6	LB99477BS	0.5	1	100	100		2.28	0.000	0.373	0.373	0.492	11/29/2018	17:31
7	J6144-07		1	100	100		1.70	0.008	0.008	0.000	0.000	11/29/2018	17:32
8	J6144-07DUP		1	100	100		1.86	0.008	0.008	0.000	0.000	11/29/2018	17:32
9	J6144-07MS	1	2	100	100		2.05	0.009	0.368	0.359	0.474	11/29/2018	17:33
10	J6144-07MSD	1	2	100	100		2.45	0.009	0.368	0.359	0.474	11/29/2018	17:33
11	CCV2	0.5	1	100	100		2.27	0.000	0.377	0.377	0.498	11/29/2018	17:34
12	CCB2		1	100	100		2.16	0.000	0.000	0.000	0.000	11/29/2018	17:34

LB99477

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Hex112918-3

WorkList ID : 119738

Date : 11/29/2018 4:23:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/30/2018	Water	J6144-07	Hexavalent Chromium	Cool 4 deg C	NEWY17	L63	14B-(1-4)-COMP	11/29/2018	SM3500-Cr B

Date/Time 11/29/18 14:00
 Received by: TSB
 Relinquished by: [Signature]

Date/Time 11/29/18 17:45
 Received by: [Signature]
 Relinquished by: [Signature]