



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 12/4/2018

OVENTEMP IN Celsius(°C): 108
Time IN: 17:45
In Date: 11/29/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 11/30/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M Sc-1

QC:LB99478

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J5761-01	AU-06-112918	1	1.13	9.85	8.5	84.5
J5761-03	AU-05-112918	2	1.13	9.57	8.57	88.2
J5762-07	BU-02-112918	3	1.13	9.98	8.22	80.1
J5765-05	NB-07-112918	4	1.11	9.67	8.61	87.6
J5770-13	OR-03-112918-A	5	1.13	9.64	8.65	88.4
J5770-15	OR-03-112918-B	6	1.12	9.93	8.96	89.0
J5771-07	EO-02-112918-A	7	1.14	9.72	8.6	86.9
J5771-09	EO-02-112918-B	8	1.13	9.77	8.8	88.8
J5771-11	EO-02-112918-C	9	1.12	9.98	8.83	87.0
J6003-22	NON-UST-03R-C	10	1.12	9.96	8.98	88.9
J6125-01	FK-GF-02-028-A	11	1.13	9.53	8.1	83.0
J6125-02	FK-GF-02-028-A	12	1.13	9.96	8.48	83.2
J6125-03	FK-GF-02-028-B	13	1.1	9.81	8.55	85.5
J6125-04	FK-GF-02-028-B	14	1.14	9.71	8.29	83.4
J6133-01	EXTERIOR-WINDOW-CAULK	15	1.00	2.00	2.00	100.0
J6139-01	S49-0483	16	1.12	9.66	9.01	92.4
J6139-02	J6139-01MS	17	1.12	9.66	9.01	92.4
J6139-03	J6139-01MSD	18	1.12	9.66	9.01	92.4
J6141-01	GD	19	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$