



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB99485

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.3	4.01	11/29/2018	17:30
2	CAL2	1	Water			21.4	7.00	11/29/2018	17:32
3	CAL3	1	Water			21.3	10.04	11/29/2018	17:34
4	ICV	1	Water			21.2	7.01	11/29/2018	17:36
5	CCV1	1	Water			21.3	2.00	11/29/2018	17:38
6	J5761-02	1	Solid	20.02	20	21.0	6.85	11/29/2018	17:40
7	J5761-02DUP	1	Solid	20.03	20	21.1	6.86	11/29/2018	17:42
8	J5761-04	1	Solid	20.02	20	20.9	6.50	11/29/2018	17:44
9	J5762-08	1	Solid	20.04	20	21.0	7.45	11/29/2018	17:46
10	J5765-06	1	Solid	20.04	20	21.0	7.87	11/29/2018	17:48
11	J5770-14	1	Solid	20.02	20	20.4	7.50	11/29/2018	17:50
12	J5770-16	1	Solid	20.03	20	20.6	7.65	11/29/2018	17:52
13	J5771-08	1	Solid	20.02	20	20.2	8.69	11/29/2018	17:54
14	J5771-10	1	Solid	20.03	20	20.2	9.09	11/29/2018	17:56
15	J5771-12	1	Solid	20.04	20	20.6	8.55	11/29/2018	17:58
16	CCV2	1	Water			21.2	12.00	11/29/2018	18:00

WORKLIST(Hardcopy Internal Chain)

LB 99485

WorkList Name : corrosivity112918

WorkList ID : 119736

Date : 11/29/2018 4:09:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/29/2018	Solid	J5761-02	Corrosivity	Cool 4 deg C	PSEG05	L31	AU-06-112918	11/29/2018	9045D
11/29/2018	Solid	J5761-04	Corrosivity	Cool 4 deg C	PSEG05	L31	AU-05-112918	11/29/2018	9045D
11/29/2018	Solid	J5762-08	Corrosivity	Cool 4 deg C	PSEG05		BU-02-112918	11/29/2018	9045D
11/29/2018	Solid	J5765-06	Corrosivity	Cool 4 deg C	PSEG05		NB-07-112918	11/29/2018	9045D
11/29/2018	Solid	J5770-14	Corrosivity	Cool 4 deg C	PSEG05		OR-03-112918-A	11/29/2018	9045D
11/29/2018	Solid	J5770-16	Corrosivity	Cool 4 deg C	PSEG05		OR-03-112918-B	11/29/2018	9045D
11/29/2018	Solid	J5771-08	Corrosivity	Cool 4 deg C	PSEG05		EO-02-112918-A	11/29/2018	9045D
11/29/2018	Solid	J5771-10	Corrosivity	Cool 4 deg C	PSEG05		EO-02-112918-B	11/29/2018	9045D
11/29/2018	Solid	J5771-12	Corrosivity	Cool 4 deg C	PSEG05		EO-02-112918-C	11/29/2018	9045D

11/29/18
 Date/Time 11/29/18 5:20
 Received by: SP
 Relinquished by: CP

11/29/18 6:05
 Date/Time 11/29/18 6:05
 Received by: SP
 Relinquished by: SP