



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB99486

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.3	4.01	11/29/2018	17:30
2	CAL2	1	Water			21.4	7.00	11/29/2018	17:32
3	CAL3	1	Water			21.3	10.03	11/29/2018	17:34
4	ICV	1	Water			21.2	7.01	11/29/2018	17:36
5	CCV1	1	Water			21.3	2.00	11/29/2018	18:02
6	J6141-01	1	Solid	20.02	20	19.3	4.51	11/29/2018	18:04
7	J6141-01DUP	1	Solid	20.03	20	19.4	4.52	11/29/2018	18:06
8	CCV2	1	Water			21.3	12.00	11/29/2018	18:08

WORKLIST(Hardcopy Internal Chain)

LB 99486

WorkList Name : pH112918

WorkList ID : 119737

Date : 11/29/2018 4:09:39 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/29/2018	Solid	J6141-01	pH	Cool 4 deg C	PSEG01	L31	GD	11/29/2018	9041A

Date/Time 11/29/18 5:20
 Received by: SP
 Relinquished by: CP

Date/Time 11/29/18 6:10
 Received by: CP
 Relinquished by: SP