

LB99487


7470A

INSTRUMENT ID: CY2

Linear

 μ Abs.:

38047



A= 0.0000e+000

B= 2.6403e-004

C= 4.6054e-003

Rho= 0.9997691

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	-43	0.000	-43				
0.2	0.200	0.190	-0.010	702	0.0 %	702				
2.5	2.500	2.500	0.000	9452	0.0 %	9452				
5.0	5.000	5.114	0.114	19353	0.0 %	19353				
7.5	7.500	7.352	-0.148	27827	0.0 %	27827				
10.0	10.000	10.050	0.050	38047	0.0 %	38047				

LB99487 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Nov 2018 10:01:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	-43	-	0.0000	7470A	PPB	30 Nov 2018 10:04:10	S	Std
0.2 - 1	SP	702	-	0.2000	7470A	PPB	30 Nov 2018 10:06:18	S	Std
2.5 - 1	SP	9452	-	2.5000	7470A	PPB	30 Nov 2018 10:08:26	S	Std
5.0 - 1	SP	19353	-	5.0000	7470A	PPB	30 Nov 2018 10:10:35	S	Std
7.5 - 1	SP	27827	-	7.5000	7470A	PPB	30 Nov 2018 10:12:43	S	Std
10.0 - 1	SP	38047	-	10.0000	7470A	PPB	30 Nov 2018 10:17:28	S	Std
ICV97 - 1	ICV97	16056	4.2439	-	7470A	PPB	30 Nov 2018 10:19:59	U	SMPL
ICB97 - 1	ICB97	-149	-0.0347	-	7470A	PPB	30 Nov 2018 10:22:05	U	SMPL
CCV33 - 1	CCV33	18930	5.0027	-	7470A	PPB	30 Nov 2018 10:24:12	U	SMPL
CCB33 - 1	CCB33	-143	-0.0332	-	7470A	PPB	30 Nov 2018 10:26:18	U	SMPL
CRA - 1	CRA	705	0.1907	-	7470A	PPB	30 Nov 2018 10:28:25	U	SMPL
HighStd - 1	HighStd	38060	10.0537	-	7470A	PPB	30 Nov 2018 10:30:32	U	SMPL
ChkStd - 1	ChkStd	26225	6.9289	-	7470A	PPB	30 Nov 2018 10:32:39	U	SMPL
PB115179BL - 1	PBW	-79	-0.0163	-	7470A	PPB	30 Nov 2018 10:34:47	U	SMPL
PB115179BS - 1	LCSW	14263	3.7705	-	7470A	PPB	30 Nov 2018 10:36:56	U	SMPL
J5767-18 - 1	CL-02-112818-A	-46	-0.0075	-	7470A	PPB	30 Nov 2018 10:39:03	U	SMPL
J5767-18DUP - 1	CL-02-112818-ADUP	-19	-0.0004	-	7470A	PPB	30 Nov 2018 10:41:11	U	SMPL
J5767-18MS - 1	CL-02-112818-AMS	13408	3.5448	-	7470A	PPB	30 Nov 2018 10:43:20	U	SMPL
J5767-18MSD - 1	CL-02-112818-AMSD	13407	3.5445	-	7470A	PPB	30 Nov 2018 10:45:28	U	SMPL
J5767-20 - 1	CL-02-112818-B	14	0.0083	-	7470A	PPB	30 Nov 2018 10:47:35	U	SMPL
CCV34 - 1	CCV34	18665	4.9328	-	7470A	PPB	30 Nov 2018 10:49:41	U	SMPL
CCB34 - 1	CCB34	-126	-0.0287	-	7470A	PPB	30 Nov 2018 10:51:48	U	SMPL
J5767-22 - 1	CL-02-112818-C	63	0.0212	-	7470A	PPB	30 Nov 2018 10:53:54	U	SMPL
J5767-24 - 1	CL-02-112818-D	148	0.0437	-	7470A	PPB	30 Nov 2018 10:56:01	U	SMPL
J5768-08 - 1	OK-01-112818	168	0.0490	-	7470A	PPB	30 Nov 2018 10:58:08	U	SMPL
PB115180BL - 1	PBW	-56	-0.0102	-	7470A	PPB	30 Nov 2018 11:02:43	U	SMPL
PB115180BS - 1	LCSW	14243	3.7652	-	7470A	PPB	30 Nov 2018 11:04:50	U	SMPL
J6124-01 - 1	BW-2072-112818	9	0.0070	-	7470A	PPB	30 Nov 2018 11:06:57	U	SMPL
J6124-01DUP - 1	BW-2072-112818DUP	38	0.0146	-	7470A	PPB	30 Nov 2018 11:09:06	U	SMPL
J6124-01MS - 1	BW-2072-112818MS	13642	3.6065	-	7470A	PPB	30 Nov 2018 11:11:14	U	SMPL
J6124-01MSD - 1	BW-2072-112818MSD	13795	3.6469	-	7470A	PPB	30 Nov 2018 11:13:22	U	SMPL
CCV35 - 1	CCV35	19020	5.0265	-	7470A	PPB	30 Nov 2018 11:15:29	U	SMPL
CCB35 - 1	CCB35	-118	-0.0266	-	7470A	PPB	30 Nov 2018 11:17:36	U	SMPL
PB115179TB - 1	PB115149TB	85	0.0270	-	7470A	PPB	30 Nov 2018 11:19:44	U	SMPL
PB115180TB - 1	PB115154TB	53	0.0186	-	7470A	PPB	30 Nov 2018 11:21:51	U	SMPL
J5767-18LX5 - 1		-51	-0.0089	-	7470A	PPB	30 Nov 2018 11:23:58	U	SMPL
J5767-18A - 1		15572	4.1161	-	7470A	PPB	30 Nov 2018 11:26:06	U	SMPL
J6124-01LX5 - 1		-90	-0.0192	-	7470A	PPB	30 Nov 2018 11:28:12	U	SMPL
J6124-01A - 1		15078	3.9857	-	7470A	PPB	30 Nov 2018 11:30:20	U	SMPL
CCV36 - 1	CCV36	18908	4.9969	-	7470A	PPB	30 Nov 2018 11:32:27	U	SMPL
CCB36 - 1	CCB36	-130	-0.0297	-	7470A	PPB	30 Nov 2018 11:34:35	U	SMPL