



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 11/30/2018

TEMP1 IN: 104 11/30/2018 07:10 TEMP OUT: 104 11/30/2018 08:10

TEMP2 IN: 104 11/30/2018 08:35 TEMP OUT: 104 11/30/2018 09:36

TEMP3 IN: 104 11/30/2018 11:10 TEMP OUT: 103 11/30/2018 12:35

TEMP4 IN: 103 11/30/2018 13:20 TEMP OUT: 103 11/30/2018 15:20

Run Number: LB99488

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99488BL	LB99488BL	2.6998	2.6998	100	2.6998	2.6998	2.6998	0.0000	0
2	LB99488BS	LB99488BS	3.1025	3.1025	100	3.1125	3.1571	3.1571	0.0546	546
3	J6126-01	COMP	1.3891	1.3891	100	1.4044	1.4044	1.4044	0.0153	153
4	J6128-05	112718-TIPC-F-D-B	1.3784	1.3784	250	1.3943	1.3943	1.3943	0.0159	63.6
5	J6128-06	112718-TIPC-F-D-S	1.3676	1.3676	250	1.3783	1.3783	1.3783	0.0107	42.8
6	J6128-07	112718-TIPC-F-U-B	1.3864	1.3864	250	1.4025	1.4025	1.4025	0.0161	64.4
7	J6128-08	112718-TIPC-F-U-S	1.4084	1.4084	250	1.4218	1.4218	1.4218	0.0134	53.6
8	J6129-01	112818-TIPC-F-D-B	1.3915	1.3915	250	1.4049	1.4049	1.4049	0.0134	53.6
9	J6129-02	112818-TIPC-F-D-S	1.3792	1.3792	250	1.3934	1.3934	1.3934	0.0142	56.8
10	J6129-03	112818-TIPC-F-U-B	1.3721	1.3721	300	1.3896	1.3896	1.3896	0.0175	58.3
11	J6129-04	112818-TIPC-F-U-S	1.3862	1.3862	300	1.4056	1.4056	1.4056	0.0194	64.7
12	J6130-01	STORM-DRAIN-1	1.3624	1.3624	800	1.3722	1.3722	1.3722	0.0098	12.3
13	J6130-02	STORM-DRAIN-2	1.3867	1.3867	400	1.3892	1.3892	1.3892	0.0025	6.3
14	J6130-02DUP	STORM-DRAIN-2DUP	1.3721	1.3721	400	1.3747	1.3747	1.3747	0.0026	6.5
15	J6144-01	14B-1	1.3729	1.3729	10	1.4336	1.4336	1.4336	0.0607	6070

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 11/30/2018

TEMP1 IN: 104 11/30/2018 07:10 TEMP OUT: 104 11/30/2018 08:10

TEMP2 IN: 104 11/30/2018 08:35 TEMP OUT: 104 11/30/2018 09:36

TEMP3 IN: 104 11/30/2018 11:10 TEMP OUT: 103 11/30/2018 12:35

TEMP4 IN: 103 11/30/2018 13:20 TEMP OUT: 103 11/30/2018 15:20

Run Number: LB99488

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J6144-02	14B-2	1.3771	1.3771	10	1.4496	1.4496	1.4496	0.0725	7250
17	J6144-03	14B-3	1.3803	1.3803	10	1.4598	1.4598	1.4598	0.0795	7950
18	J6144-04	14B-4	1.3854	1.3854	10	1.4806	1.4806	1.4806	0.0952	9520
19	J6145-01	SPDES-1	1.3856	1.3856	800	1.3859	1.3859	1.3859	0.0003	0.4

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

LB 99488

WorkList Name : tss j6145

WorkList ID : 119768

Date : 11/30/2018 9:30:01 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/05/2018	Water	J6126-01 ✓	TSS	Cool 4 deg C	ARAM01	L13	COMP	11/28/2018	SM2540 D
11/30/2018	Water	J6128-05 -C	TSS	Cool 4 deg C	HDR103	L42	112718-TIPC-F-D-B	11/27/2018	SM2540 D
11/30/2018	Water	J6128-06 -C	TSS	Cool 4 deg C	HDR103	L42	112718-TIPC-F-D-S	11/27/2018	SM2540 D
11/30/2018	Water	J6128-07 -C	TSS	Cool 4 deg C	HDR103	L42	112718-TIPC-F-U-B	11/27/2018	SM2540 D
11/30/2018	Water	J6128-08 -C	TSS	Cool 4 deg C	HDR103	L42	112718-TIPC-F-U-S	11/27/2018	SM2540 D
11/30/2018	Water	J6129-01 -C	TSS	Cool 4 deg C	HDR103	L31	112818-TIPC-F-D-B	11/28/2018	SM2540 D
11/30/2018	Water	J6129-02 -C	TSS	Cool 4 deg C	HDR103	L31	112818-TIPC-F-D-S	11/28/2018	SM2540 D
11/30/2018	Water	J6129-03 -C	TSS	Cool 4 deg C	HDR103	L31	112818-TIPC-F-U-B	11/28/2018	SM2540 D
11/30/2018	Water	J6129-04 -C	TSS	Cool 4 deg C	HDR103	L31	112818-TIPC-F-U-S	11/28/2018	SM2540 D
12/05/2018	Water	J6130-01 -C	TSS	Cool 4 deg C	KWAR01	L41	STORM-DRAIN-1	11/28/2018	SM2540 D
12/05/2018	Water	J6130-02 -C	TSS	Cool 4 deg C	KWAR01	L41	STORM-DRAIN-2	11/28/2018	SM2540 D
12/06/2018	Water	J6144-01 -C	TSS	Cool 4 deg C	NEWY17	L63	14B-1	11/29/2018	SM2540 D
12/06/2018	Water	J6144-02 -C	TSS	Cool 4 deg C	NEWY17	L63	14B-2	11/29/2018	SM2540 D
12/06/2018	Water	J6144-03 -C	TSS	Cool 4 deg C	NEWY17	L63	14B-3	11/29/2018	SM2540 D
12/06/2018	Water	J6144-04 -C	TSS	Cool 4 deg C	NEWY17	L63	14B-4	11/29/2018	SM2540 D
11/30/2018	Water	J6145-01 -D	TSS	Cool 4 deg C	SKAN06	L52	SPDES-1	11/29/2018	SM2540 D

Date/Time 11-30-18 9:35 AM
 Received by: J0
 Relinquished by: [Signature]

Date/Time 11-30-18 12:00 PM
 Received by: [Signature]
 Relinquished by: J0