

LB99492


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35741



A= 0.0000e+000

B= 2.8060e-004

C= 1.2719e-002

Rho= 0.9999584

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	-53	0.000	-53				
0.05	0.050	0.048	-0.002	124	0.0 %	124				
0.2	0.200	0.195	-0.005	649	0.0 %	649				
2.5	2.500	2.535	0.035	8989	0.0 %	8989				
5.0	5.000	5.004	0.004	17787	0.0 %	17787				
7.5	7.500	7.429	-0.071	26431	0.0 %	26431				
10.0	10.000	10.042	0.042	35741	0.0 %	35741				

LB99492 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 30 Nov 2018 11:43:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-53	-	0.0000	245.1	PPB	30 Nov 2018 11:45:03	S	Std
0.05 - 1	50.5	124	-	0.0500	245.1	PPB	30 Nov 2018 11:47:21	S	Std
0.2 - 1	50.2	649	-	0.2000	245.1	PPB	30 Nov 2018 11:49:39	S	Std
2.5 - 1	52.5	8989	-	2.5000	245.1	PPB	30 Nov 2018 11:51:57	S	Std
5.0 - 1	55	17787	-	5.0000	245.1	PPB	30 Nov 2018 11:54:14	S	Std
7.5 - 1	57.5	26431	-	7.5000	245.1	PPB	30 Nov 2018 11:56:30	S	Std
10.0 - 1	510	35741	-	10.0000	245.1	PPB	30 Nov 2018 11:58:48	S	Std
ICV98 - 1	ICV98	13722	3.8631	-245.1		PPB	30 Nov 2018 12:01:50	U	SMPL
ICB98 - 1	ICB98	-171	-0.0353	-245.1		PPB	30 Nov 2018 12:04:06	U	SMPL
CCV37 - 1	CCV37	18103	5.0924	-245.1		PPB	30 Nov 2018 12:06:22	U	SMPL
CCB37 - 1	CCB37	-168	-0.0344	-245.1		PPB	30 Nov 2018 12:08:38	U	SMPL
CRA - 1	CRA	128	0.0486	-245.1		PPB	30 Nov 2018 12:10:55	U	SMPL
HighStd - 1	HighStd	35827	10.0658	-245.1		PPB	30 Nov 2018 12:13:11	U	SMPL
ChkStd - 1	ChkStd	23558	6.6231	-245.1		PPB	30 Nov 2018 12:18:03	U	SMPL
PB115187BL - 1	PBW	24	0.0195	-245.1		PPB	30 Nov 2018 12:20:19	U	SMPL
PB115187BS - 1	LCSW	13184	3.7122	-245.1		PPB	30 Nov 2018 12:22:38	U	SMPL
J6086-07 - 1	14B-(1-4)-COMP	1157	0.3374	-245.1		PPB	30 Nov 2018 12:24:55	U	SMPL
J6120-07 - 1	14B-(1-4)-COMP	828	0.2451	-245.1		PPB	30 Nov 2018 12:27:12	U	SMPL
J6128-01 - 1	112718-DBRI-E-D-B	32	0.0217	-245.1		PPB	30 Nov 2018 12:29:29	U	SMPL
J6128-02 - 1	112718-DBRI-E-D-S	80	0.0352	-245.1		PPB	30 Nov 2018 12:31:47	U	SMPL
J6128-03 - 1	112718-DBRI-E-U-B	384	0.1205	-245.1		PPB	30 Nov 2018 12:34:03	U	SMPL
CCV38 - 1	CCV38	17869	5.0268	-245.1		PPB	30 Nov 2018 12:36:19	U	SMPL
CCB38 - 1	CCB38	-148	-0.0288	-245.1		PPB	30 Nov 2018 12:38:34	U	SMPL
J6128-04 - 1	112718-DBRI-E-U-S	264	0.0868	-245.1		PPB	30 Nov 2018 12:40:51	U	SMPL
J6128-05 - 1	112718-TIPC-F-D-B	171	0.0607	-245.1		PPB	30 Nov 2018 12:43:07	U	SMPL
J6128-06 - 1	112718-TIPC-F-D-S	136	0.0509	-245.1		PPB	30 Nov 2018 12:45:22	U	SMPL
J6128-07 - 1	112718-TIPC-F-U-B	137	0.0512	-245.1		PPB	30 Nov 2018 12:47:39	U	SMPL
J6128-08 - 1	112718-TIPC-F-U-S	135	0.0506	-245.1		PPB	30 Nov 2018 12:49:55	U	SMPL
J6128-09 - 1	112718-CHDE-F-D-B	25	0.0197	-245.1		PPB	30 Nov 2018 12:52:11	U	SMPL
J6128-10 - 1	112718-CHDE-F-D-M	38	0.0234	-245.1		PPB	30 Nov 2018 12:54:29	U	SMPL
J6128-11 - 1	112718-CHDE-F-D-S	62	0.0301	-245.1		PPB	30 Nov 2018 12:56:46	U	SMPL
J6128-12 - 1	112718-CHDE-F-U-B	88	0.0374	-245.1		PPB	30 Nov 2018 12:59:04	U	SMPL
J6128-13 - 1	112718-CHDE-F-U-M	86	0.0369	-245.1		PPB	30 Nov 2018 13:01:20	U	SMPL
CCV39 - 1	CCV39	17901	5.0358	-245.1		PPB	30 Nov 2018 13:03:37	U	SMPL
CCB39 - 1	CCB39	-161	-0.0325	-245.1		PPB	30 Nov 2018 13:05:53	U	SMPL
J6128-14 - 1	112718-CHDE-F-U-S	51	0.0270	-245.1		PPB	30 Nov 2018 13:08:10	U	SMPL
J6128-14DUP - 1	112718-CHDE-F-U-SDUP	-58	-0.0036	-245.1		PPB	30 Nov 2018 13:10:25	U	SMPL
J6128-14MS - 1	112718-CHDE-F-U-SMS	14404	4.0545	-245.1		PPB	30 Nov 2018 13:12:42	U	SMPL
J6128-14MSD - 1	112718-CHDE-F-U-SMSD	14318	4.0304	-245.1		PPB	30 Nov 2018 13:14:58	U	SMPL
PB115188BL - 1	PBW	329	0.1050	-245.1		PPB	30 Nov 2018 13:17:15	U	SMPL
PB115188BS - 1	LCSW	13215	3.7209	-245.1		PPB	30 Nov 2018 13:19:32	U	SMPL
J6129-01 - 1	112818-TIPC-F-D-B	385	0.1208	-245.1		PPB	30 Nov 2018 13:21:49	U	SMPL
J6129-02 - 1	112818-TIPC-F-D-S	-13	0.0091	-245.1		PPB	30 Nov 2018 13:24:07	U	SMPL
J6129-03 - 1	112818-TIPC-F-U-B	5	0.0141	-245.1		PPB	30 Nov 2018 13:26:25	U	SMPL
J6129-04 - 1	112818-TIPC-F-U-S	-16	0.0082	-245.1		PPB	30 Nov 2018 13:28:41	U	SMPL
CCV40 - 1	CCV40	17603	4.9521	-245.1		PPB	30 Nov 2018 13:30:58	U	SMPL
CCB40 - 1	CCB40	-159	-0.0319	-245.1		PPB	30 Nov 2018 13:33:14	U	SMPL
J6129-05 - 1	112818-DBRI-F-D-B	-57	-0.0033	-245.1		PPB	30 Nov 2018 13:35:31	U	SMPL
J6129-06 - 1	112818-DBRI-F-D-S	-43	0.0007	-245.1		PPB	30 Nov 2018 13:37:48	U	SMPL
J6129-07 - 1	112818-DBRI-F-U-B	-50	-0.0013	-245.1		PPB	30 Nov 2018 13:40:04	U	SMPL
J6129-08 - 1	112818-DBRI-F-U-S	424	0.1317	-245.1		PPB	30 Nov 2018 13:42:21	U	SMPL
J6129-09 - 1	112818-CHDE-F-D-B	475	0.1460	-245.1		PPB	30 Nov 2018 13:44:37	U	SMPL
J6129-10 - 1	112818-CHDE-F-D-M	400	0.1250	-245.1		PPB	30 Nov 2018 13:46:54	U	SMPL
J6129-11 - 1	112818-CHDE-F-D-S	416	0.1294	-245.1		PPB	30 Nov 2018 13:49:11	U	SMPL
J6129-12 - 1	112818-CHDE-F-U-B	410	0.1278	-245.1		PPB	30 Nov 2018 13:51:29	U	SMPL
J6129-13 - 1	112818-CHDE-F-U-M	437	0.1353	-245.1		PPB	30 Nov 2018 13:53:47	U	SMPL
J6129-14 - 1	112818-CHDE-F-U-S	-102	-0.0159	-245.1		PPB	30 Nov 2018 13:56:03	U	SMPL
CCV41 - 1	CCV41	17742	4.9911	-245.1		PPB	30 Nov 2018 13:58:19	U	SMPL
CCB41 - 1	CCB41	-158	-0.0316	-245.1		PPB	30 Nov 2018 14:00:36	U	SMPL
J6129-14DUP - 1	112818-CHDE-F-U-SDUP	-72	-0.0075	-245.1		PPB	30 Nov 2018 14:02:52	U	SMPL
J6129-14MS - 1	112818-CHDE-F-U-SMS	14024	3.9479	-245.1		PPB	30 Nov 2018 14:05:09	U	SMPL
J6129-14MSD - 1	112818-CHDE-F-U-SMSD	14441	4.0649	-245.1		PPB	30 Nov 2018 14:07:25	U	SMPL
J6144-07 - 1	14B-(1-4)-COMP	599	0.1808	-245.1		PPB	30 Nov 2018 14:09:42	U	SMPL
J6128-14LX5 - 1		-98	-0.0148	-245.1		PPB	30 Nov 2018 14:11:59	U	SMPL
J6128-14A - 1		16809	4.7293	-245.1		PPB	30 Nov 2018 14:14:16	U	SMPL
J6129-14LX5 - 1		-133	-0.0246	-245.1		PPB	30 Nov 2018 14:16:33	U	SMPL
J6129-14A - 1		14752	4.1522	-245.1		PPB	30 Nov 2018 14:18:50	U	SMPL
CCV42 - 1	CCV42	17886	5.0316	-245.1		PPB	30 Nov 2018 14:21:08	U	SMPL
CCB42 - 1	CCB42	-156	-0.0311	-245.1		PPB	30 Nov 2018 14:23:24	U	SMPL