


7471B

LB99499
INSTRUMENT ID: CVL

Linear

 μ Abs.:

55598



A= 0.0000e+000

B= 1.8125e-004

C= -2.6061e-002

Rho= 0.9996289

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.032	-0.032	-33	0.000	-33				
0.2	0.200	0.165	-0.035	1053	0.0 %	1053				
2.5	2.500	2.573	0.073	14340	0.0 %	14340				
5.0	5.000	5.128	0.128	28437	0.0 %	28437				
7.5	7.500	7.315	-0.185	40506	0.0 %	40506				
10.0	10.000	10.051	0.051	55598	0.0 %	55598				

LB99499 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Nov 2018 15:38:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	500	-33	-	0.0000	7471B	PPB	30 Nov 2018 15:44:23	S	Std
0.2 - 1	502	1053	-	0.2000	7471B	PPB	30 Nov 2018 15:46:41	S	Std
2.5 - 1	525	14340	-	2.5000	7471B	PPB	30 Nov 2018 15:48:59	S	Std
5.0 - 1	550	28437	-	5.0000	7471B	PPB	30 Nov 2018 15:51:17	S	Std
7.5 - 1	575	40506	-	7.5000	7471B	PPB	30 Nov 2018 15:56:12	S	Std
10.0 - 1	5100	55598	-	10.0000	7471B	PPB	30 Nov 2018 15:58:28	S	Std
ICV99 - 1	ICV99	22112	3.9816	-7471B	PPB	PPB	30 Nov 2018 16:01:21	U	SMPL
ICB99 - 1	ICB99	-133	-0.0502	-7471B	PPB	PPB	30 Nov 2018 16:03:37	U	SMPL
CCV43 - 1	CCV43	27847	5.0211	-7471B	PPB	PPB	30 Nov 2018 16:05:53	U	SMPL
CCB43 - 1	CCB43	-134	-0.0503	-7471B	PPB	PPB	30 Nov 2018 16:08:10	U	SMPL
CRA - 1	CRA	1182	0.1882	-7471B	PPB	PPB	30 Nov 2018 16:16:00	U	SMPL
HighStd - 1	HighStd	55650	10.0602	-7471B	PPB	PPB	30 Nov 2018 16:18:15	U	SMPL
ChkStd - 1	ChkStd	38563	6.9633	-7471B	PPB	PPB	30 Nov 2018 16:22:37	U	SMPL
PB115192BL - 1	PBS	-206	-0.0634	-7471B	PPB	PPB	30 Nov 2018 16:24:53	U	SMPL
PB115192BS - 1	LCSS	19463	3.5015	-7471B	PPB	PPB	30 Nov 2018 16:27:12	U	SMPL
J5761-01 - 1	AU-06-112918	51464	9.3015	-7471B	PPB	PPB	30 Nov 2018 16:29:29	U	SMPL
J5761-03 - 1	AU-05-112918	41506	7.4967	-7471B	PPB	PPB	30 Nov 2018 16:31:47	U	SMPL
J5762-07 - 1	BU-02-112918	1768	0.2944	-7471B	PPB	PPB	30 Nov 2018 16:34:06	U	SMPL
J5765-05 - 1	NB-07-112918	726	0.1055	-7471B	PPB	PPB	30 Nov 2018 16:36:26	U	SMPL
CCV44 - 1	CCV44	27343	4.9297	-7471B	PPB	PPB	30 Nov 2018 16:38:43	U	SMPL
CCB44 - 1	CCB44	-184	-0.0594	-7471B	PPB	PPB	30 Nov 2018 16:40:58	U	SMPL
J5770-13 - 1	OR-03-112918-A	4252	0.7446	-7471B	PPB	PPB	30 Nov 2018 16:43:15	U	SMPL
J5770-13DUP - 1	OR-03-112918-ADUP	4216	0.7381	-7471B	PPB	PPB	30 Nov 2018 16:45:32	U	SMPL
J5770-13MS - 1	OR-03-112918-AMS	23862	4.2988	-7471B	PPB	PPB	30 Nov 2018 16:47:48	U	SMPL
J5770-13MSD - 1	OR-03-112918-AMSD	23838	4.2945	-7471B	PPB	PPB	30 Nov 2018 16:50:04	U	SMPL
J5770-15 - 1	OR-03-112918-B	5056	0.8903	-7471B	PPB	PPB	30 Nov 2018 16:52:21	U	SMPL
J5771-07 - 1	EO-02-112918-A	4146	0.7254	-7471B	PPB	PPB	30 Nov 2018 16:54:39	U	SMPL
J5771-09 - 1	EO-02-112918-B	7076	1.2564	-7471B	PPB	PPB	30 Nov 2018 16:56:56	U	SMPL
J5771-11 - 1	EO-02-112918-C	5843	1.0330	-7471B	PPB	PPB	30 Nov 2018 16:59:14	U	SMPL
J6125-01 - 1	FK-GF-02-028-A	7333	1.3030	-7471B	PPB	PPB	30 Nov 2018 17:01:31	U	SMPL
J6125-03 - 1	FK-GF-02-028-B	7813	1.3900	-7471B	PPB	PPB	30 Nov 2018 17:03:49	U	SMPL
CCV45 - 1	CCV45	27887	5.0283	-7471B	PPB	PPB	30 Nov 2018 17:06:05	U	SMPL
CCB45 - 1	CCB45	-178	-0.0583	-7471B	PPB	PPB	30 Nov 2018 17:08:22	U	SMPL
J5770-13LX5 - 1		709	0.1024	-7471B	PPB	PPB	30 Nov 2018 17:10:39	U	SMPL
J5770-13A - 1		24940	4.4942	-7471B	PPB	PPB	30 Nov 2018 17:12:56	U	SMPL
CCV46 - 1	CCV46	27650	4.9854	-7471B	PPB	PPB	30 Nov 2018 17:15:12	U	SMPL
CCB46 - 1	CCB46	-176	-0.0580	-7471B	PPB	PPB	30 Nov 2018 17:17:30	U	SMPL