



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99501

Slope : 99.1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	11/30/2018	14:10
2	CAL2	1	Water			21.5	7.00	11/30/2018	14:12
3	CAL3	1	Water			21.5	10.04	11/30/2018	14:14
4	ICV	1	Water			21.4	7.01	11/30/2018	14:16
5	CCV1	1	Water			21.5	2.00	11/30/2018	14:18
6	J6153-01	1	Water			16.2	8.73	11/30/2018	14:20
7	J6153-01DUP	1	Water			16.3	8.74	11/30/2018	14:22
8	J6153-02	1	Water			16.2	8.62	11/30/2018	14:24
9	J6153-03	1	Water			16.3	8.49	11/30/2018	14:26
10	J6153-04	1	Water			16.7	8.14	11/30/2018	14:28
11	J6153-05	1	Water			16.9	7.93	11/30/2018	14:30
12	CCV2	1	Water			21.5	12.00	11/30/2018	14:32

WORKLIST(Hardcopy Internal Chain)

LB 99501

WorkList Name : pH113018

WorkList ID : 119782

Date : 11/30/2018 1:38:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/29/2018	Water	J6153-01	pH	Cool 4 deg C	TULL02	L22	MH-F	11/29/2018	9040C
11/29/2018	Water	J6153-02	pH	Cool 4 deg C	TULL02	L22	MH-G	11/29/2018	9040C
11/29/2018	Water	J6153-03	pH	Cool 4 deg C	TULL02	L22	MH-J	11/29/2018	9040C
11/29/2018	Water	J6153-04	pH	Cool 4 deg C	TULL02	L22	MH-H	11/29/2018	9040C
11/29/2018	Water	J6153-05	pH	Cool 4 deg C	TULL02	L22	MH-9.41	11/29/2018	9040C

Date/Time 11/30/18 2:00

Received by: _____

Relinquished by: _____

Date/Time 11/30/18 2:40

Received by: _____

Relinquished by: _____