



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB99519

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.4	4.01	11/30/2018	17:30
2	CAL2	1	Water			21.4	7.00	11/30/2018	17:32
3	CAL3	1	Water			21.2	10.03	11/30/2018	17:34
4	ICV	1	Water			21.3	7.01	11/30/2018	17:36
5	CCV1	1	Water			21.3	2.00	11/30/2018	17:38
6	J5764-22	1	Solid	20.02	20	20.1	8.43	11/30/2018	17:40
7	J5764-22DUP	1	Solid	20.04	20	20.2	8.44	11/30/2018	17:42
8	J5764-24	1	Solid	20.02	20	20.0	8.71	11/30/2018	17:44
9	J5764-26	1	Solid	20.02	20	20.0	9.04	11/30/2018	17:46
10	J5764-28	1	Solid	20.03	20	19.9	8.92	11/30/2018	17:48
11	J5764-30	1	Solid	20.04	20	19.8	8.80	11/30/2018	17:50
12	J5775-04	1	Solid	20.02	20	19.9	9.41	11/30/2018	17:52
13	J6148-02	1	Solid	20.03	20	21.0	8.92	11/30/2018	17:54
14	J6148-04	1	Solid	20.04	20	21.0	8.96	11/30/2018	17:56
15	J6149-01	1	Solid	20.04	20	19.5	7.05	11/30/2018	17:58
16	CCV2	1	Water			21.2	12.00	11/30/2018	18:00
17	J6149-02	1	Solid	20.02	20	19.2	7.38	11/30/2018	18:02
18	J6149-03	1	Solid	20.02	20	18.0	6.86	11/30/2018	18:04
19	J6149-03DUP	1	Solid	20.03	20	18.1	6.87	11/30/2018	18:06
20	J6149-04	1	Solid	20.03	20	19.3	8.07	11/30/2018	18:08

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	J6150-01	1	Solid	20.02	20	21.0	5.00	11/30/2018	18:10
22	J6159-01	1	Solid	20.02	20	20.3	7.73	11/30/2018	18:12
23	J6159-02	1	Solid	20.03	20	20.4	7.94	11/30/2018	18:14
24	J6159-03	1	Solid	20.03	20	20.2	6.95	11/30/2018	18:16
25	J6159-04	1	Solid	20.04	20	19.9	6.96	11/30/2018	18:20
26	J6160-04	1	Solid	20.02	20	20.7	8.71	11/30/2018	18:22
27	CCV3	1	Water			21.3	2.00	11/30/2018	18:24

WORKLIST(Hardcopy Internal Chain)

1399519

WorkList Name : corrosivity113018

WorkList ID : 119788

Date : 11/30/2018 3:38:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/30/2018	Solid	J5764-22	Corrosivity	Cool 4 deg C	PSEG05		JC-02-113018-A	11/30/2018	9045D
11/30/2018	Solid	J5764-24	Corrosivity	Cool 4 deg C	PSEG05		JC-02-113018-B	11/30/2018	9045D
11/30/2018	Solid	J5764-26	Corrosivity	Cool 4 deg C	PSEG05		JC-02-113018-C	11/30/2018	9045D
11/30/2018	Solid	J5764-28	Corrosivity	Cool 4 deg C	PSEG05		JC-02-113018-D	11/30/2018	9045D
11/30/2018	Solid	J5764-30	Corrosivity	Cool 4 deg C	PSEG05		JC-02-113018-E	11/30/2018	9045D
11/30/2018	Solid	J5775-04	Corrosivity	Cool 4 deg C	PSEG05		TR-04-113018	11/30/2018	9045D
11/27/2018	Solid	J6148-02	Corrosivity	Cool 4 deg C	TULL02	L13	TP-MH-F	11/27/2018	9045D
11/27/2018	Solid	J6148-04	Corrosivity	Cool 4 deg C	TULL02	L13	TP-MH-J	11/27/2018	9045D
11/27/2018	Solid	J6149-01	Corrosivity	Cool 4 deg C	CLEA11	L32	ENGINE-WASH	11/27/2018	9045D
11/27/2018	Solid	J6149-02	Corrosivity	Cool 4 deg C	CLEA11	L32	OWS-2	11/27/2018	9045D
11/27/2018	Solid	J6149-03	Corrosivity	Cool 4 deg C	CLEA11	L32	BUS-WASH	11/27/2018	9045D
11/27/2018	Solid	J6149-04	Corrosivity	Cool 4 deg C	CLEA11	L32	RECLAIMER-PIT	11/27/2018	9045D
11/30/2018	Solid	J6150-01	Corrosivity	Cool 4 deg C	PSEG01	L31	WASTE-OIL-TOTES	11/30/2018	9045D
11/29/2018	Solid	J6159-01	Corrosivity	Cool 4 deg C	USAR03	L52	CB-1129-S-TCLP-GRID19	11/29/2018	9045D
11/29/2018	Solid	J6159-02	Corrosivity	Cool 4 deg C	USAR03	L52	CB-1129-S-TCLP-GRID26-1	11/29/2018	9045D
11/29/2018	Solid	J6159-03	Corrosivity	Cool 4 deg C	USAR03	L52	CB-1129-S-TCLP-GRID26-2	11/29/2018	9045D
11/29/2018	Solid	J6159-04	Corrosivity	Cool 4 deg C	USAR03	L52	CB-1129-S-TCLP-GRID27	11/29/2018	9045D
11/30/2018	Solid	J6160-04	Corrosivity	Cool 4 deg C	PSEG01	L31	TEST-PIT	11/30/2018	9045D

Date/Time 11/30/18 5:20
 Received by: SP
 Relinquished by: CP

Date/Time 11/30/18 6:30
 Received by: CP
 Relinquished by: SP