

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/03/2018

TEMP1 IN: 103 12/03/2018 07:00 TEMP OUT: 104 12/03/2018 08:10

TEMP2 IN: 104 12/03/2018 09:00 TEMP OUT: 104 12/03/2018 10:10

TEMP3 IN: 105 12/03/2018 11:30 TEMP OUT: 105 12/03/2018 13:05

TEMP4 IN: 105 12/03/2018 14:00 TEMP OUT: 105 12/03/2018 15:40

Run Number: LB99520

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99520BL	LB99520BL	1.4000	1.4000	100	1.4000	1.4000	1.4000	0.0000	0
2	LB99520BS	LB99520BS	1.3852	1.3852	100	1.4360	1.4360	1.4360	0.0508	508
3	J6128-01	112718-DBRI-E-D-B	1.3900	1.3900	200	1.4004	1.4004	1.4004	0.0104	52
4	J6128-01DUP	112718-DBRI-E-D-BDUP	1.3982	1.3982	200	1.4085	1.4085	1.4085	0.0103	51.5
5	J6128-02	112718-DBRI-E-D-S	1.3851	1.3851	500	1.4080	1.4080	1.4080	0.0229	45.8
6	J6128-03	112718-DBRI-E-U-B	1.4171	1.4171	300	1.4358	1.4358	1.4358	0.0187	62.3
7	J6128-04	112718-DBRI-E-U-S	1.4079	1.4079	300	1.4224	1.4224	1.4224	0.0145	48.3
8	J6128-09	112718-CHDE-F-D-B	1.4094	1.4094	400	1.4233	1.4233	1.4233	0.0139	34.8
9	J6128-10	112718-CHDE-F-D-M	1.4091	1.4091	400	1.4243	1.4243	1.4243	0.0152	38
10	J6128-11	112718-CHDE-F-D-S	1.3982	1.3982	400	1.4133	1.4133	1.4133	0.0151	37.8
11	J6128-12	112718-CHDE-F-U-B	1.4039	1.4039	300	1.4207	1.4207	1.4207	0.0168	56
12	J6128-13	112718-CHDE-F-U-M	1.3917	1.3917	400	1.4061	1.4061	1.4061	0.0144	36
13	J6128-14	112718-CHDE-F-U-S	1.4001	1.4001	400	1.4128	1.4128	1.4128	0.0127	31.8
14	J6129-05	112818-DBRI-F-D-B	1.3774	1.3774	300	1.3953	1.3953	1.3953	0.0179	59.7
15	J6129-06	112818-DBRI-F-D-S	1.4197	1.4197	300	1.4353	1.4353	1.4353	0.0156	52

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/03/2018

TEMP1 IN: 103 12/03/2018 07:00 TEMP OUT: 104 12/03/2018 08:10

TEMP2 IN: 104 12/03/2018 09:00 TEMP OUT: 104 12/03/2018 10:10

TEMP3 IN: 105 12/03/2018 11:30 TEMP OUT: 105 12/03/2018 13:05

TEMP4 IN: 105 12/03/2018 14:00 TEMP OUT: 105 12/03/2018 15:40

Run Number: LB99520

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J6129-07	112818-DBRI-F-U-B	1.3770	1.3770	300	1.3916	1.3916	1.3916	0.0146	48.7
17	J6129-08	112818-DBRI-F-U-S	1.4125	1.4125	300	1.4257	1.4257	1.4257	0.0132	44
18	J6129-09	112818-CHDE-F-D-B	1.4123	1.4123	300	1.4259	1.4259	1.4259	0.0136	45.3
19	J6129-10	112818-CHDE-F-D-M	1.3957	1.3957	300	1.4117	1.4117	1.4117	0.0160	53.3
20	J6129-11	112818-CHDE-F-D-S	1.3801	1.3801	300	1.3936	1.3936	1.3936	0.0135	45
21	J6129-12	112818-CHDE-F-U-B	1.3908	1.3908	300	1.4069	1.4069	1.4069	0.0161	53.7
22	J6129-13	112818-CHDE-F-U-M	1.3867	1.3867	300	1.4016	1.4016	1.4016	0.0149	49.7
23	J6129-14	112818-CHDE-F-U-S	1.3964	1.3964	200	1.4053	1.4053	1.4053	0.0089	44.5
24	J6129-14DUP	112818-CHDE-F-U-SDUP	1.3856	1.3856	200	1.3946	1.3946	1.3946	0.0090	45

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS120318

WorkList ID : 119823

Date : 12/3/2018 9:50:14 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/04/2018	Water	J6128-01 C	TSS	Cool 4 deg C	HDR103	L42	112718-DBRI-E-D-B	11/27/2018	SM2540 D
12/04/2018	Water	J6128-02 C	TSS	Cool 4 deg C	HDR103	L42	112718-DBRI-E-D-S	11/27/2018	SM2540 D
12/04/2018	Water	J6128-03 C	TSS	Cool 4 deg C	HDR103	L42	112718-DBRI-E-U-B	11/27/2018	SM2540 D
12/04/2018	Water	J6128-04 C	TSS	Cool 4 deg C	HDR103	L42	112718-DBRI-E-U-S	11/27/2018	SM2540 D
12/04/2018	Water	J6128-09 C	TSS	Cool 4 deg C	HDR103	L42	112718-CHDE-F-D-B	11/27/2018	SM2540 D
12/04/2018	Water	J6128-10 C	TSS	Cool 4 deg C	HDR103	L42	112718-CHDE-F-D-M	11/27/2018	SM2540 D
12/04/2018	Water	J6128-11 C	TSS	Cool 4 deg C	HDR103	L42	112718-CHDE-F-D-S	11/27/2018	SM2540 D
12/04/2018	Water	J6128-12 C	TSS	Cool 4 deg C	HDR103	L42	112718-CHDE-F-U-B	11/27/2018	SM2540 D
12/04/2018	Water	J6128-13 C	TSS	Cool 4 deg C	HDR103	L42	112718-CHDE-F-U-M	11/27/2018	SM2540 D
12/04/2018	Water	J6128-14 C	TSS	Cool 4 deg C	HDR103	L42	112718-CHDE-F-U-S	11/27/2018	SM2540 D
12/05/2018	Water	J6129-05 C	TSS	Cool 4 deg C	HDR103	L31	112818-DBRI-F-D-B	11/28/2018	SM2540 D
12/05/2018	Water	J6129-06 C	TSS	Cool 4 deg C	HDR103	L31	112818-DBRI-F-D-S	11/28/2018	SM2540 D
12/05/2018	Water	J6129-07 C	TSS	Cool 4 deg C	HDR103	L31	112818-DBRI-F-U-B	11/28/2018	SM2540 D
12/05/2018	Water	J6129-08 C	TSS	Cool 4 deg C	HDR103	L31	112818-DBRI-F-U-S	11/28/2018	SM2540 D
12/05/2018	Water	J6129-09 C	TSS	Cool 4 deg C	HDR103	L31	112818-CHDE-F-D-B	11/28/2018	SM2540 D
12/05/2018	Water	J6129-10 C	TSS	Cool 4 deg C	HDR103	L31	112818-CHDE-F-D-M	11/28/2018	SM2540 D
12/05/2018	Water	J6129-11 C	TSS	Cool 4 deg C	HDR103	L31	112818-CHDE-F-D-S	11/28/2018	SM2540 D
12/05/2018	Water	J6129-12 C	TSS	Cool 4 deg C	HDR103	L31	112818-CHDE-F-U-B	11/28/2018	SM2540 D
12/05/2018	Water	J6129-13 C	TSS	Cool 4 deg C	HDR103	L31	112818-CHDE-F-U-M	11/28/2018	SM2540 D
12/05/2018	Water	J6129-14 C	TSS	Cool 4 deg C	HDR103	L31	112818-CHDE-F-U-S	11/28/2018	SM2540 D

fp-12791048

Date/Time

12/04/18, 10:30

Received by:

AS

Relinquished by:

CS

Date/Time

12/04/18 13:00

Received by:

CS

Relinquished by:

AS