



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99527

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.8	4.01	12/03/2018	10:40
2	CAL2	1	Water			20.6	7.00	12/03/2018	10:42
3	CAL3	1	Water			20.8	10.04	12/03/2018	10:44
4	ICV	1	Water			20.7	7.01	12/03/2018	10:46
5	CCV1	1	Water			20.6	2.00	12/03/2018	10:48
6	J6183-01	1	Water			18.6	5.33	12/03/2018	10:50
7	J6183-01DUP	1	Water			18.7	5.34	12/03/2018	10:52
8	CCV2	1	Water			20.4	12.00	12/03/2018	10:54

WORKLIST(Hardcopy Internal Chain)

LB 99 527

WorkList Name : pH120318

WorkList ID : 119836

Date : 12/3/2018 10:27:19 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/17/2018	Water	J6183-01 A	pH	Cool 4 deg C	STER14	L11	JS-SYSTEM	12/29/2018	SM 4500-H B

Date/Time 12/03/18 10:30
 Received by: SP
 Relinquished by: SP

Date/Time 12/03/18 11:00
 Received by: SP
 Relinquished by: SP