

LB 99537


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53683



A= 0.0000e+000

B= 1.8620e-004

C= 1.3741e-002

Rho= 0.9999834

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	-37	0.000	-37				
0.2	0.200	0.173	-0.027	853	0.0 %	853				
2.5	2.500	2.535	0.035	13543	0.0 %	13543				
5.0	5.000	4.999	-0.001	26775	0.0 %	26775				
7.5	7.500	7.476	-0.024	40078	0.0 %	40078				
10.0	10.000	10.010	0.010	53683	0.0 %	53683				

LB99537

Method: 7471B

Operator: Admin

Date of Analysis: 03 Dec 2018 13:42:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-37	-	0.00007471B	PPB		03 Dec 2018 13:52:00	S	Std
0.2 - 1	50.2	853	-	0.20007471B	PPB		03 Dec 2018 13:54:18	S	Std
2.5 - 1	52.5	13543	-	2.50007471B	PPB		03 Dec 2018 13:56:36	S	Std
5.0 - 1	55	26775	-	5.00007471B	PPB		03 Dec 2018 13:58:54	S	Std
7.5 - 1	57.5	40078	-	7.50007471B	PPB		03 Dec 2018 14:01:11	S	Std
10.0 - 1	59	53683	-	10.00007471B	PPB		03 Dec 2018 14:06:06	S	Std
ICV02 - 1	ICV02	21450	4.0077	-7471B	PPB		03 Dec 2018 14:08:48	U	SMPL
ICB02 - 1	ICB02	-184	-0.0205	-7471B	PPB		03 Dec 2018 14:11:03	U	SMPL
CCV53 - 1	CCV53	26927	5.0276	-7471B	PPB		03 Dec 2018 14:13:20	U	SMPL
CCB53 - 1	CCB53	-190	-0.0216	-7471B	PPB		03 Dec 2018 14:15:36	U	SMPL
CRA - 1	CRA	880	0.1776	-7471B	PPB		03 Dec 2018 14:17:53	U	SMPL
HighStd - 1	HighStd	54221	10.1097	-7471B	PPB		03 Dec 2018 14:20:09	U	SMPL
ChkStd - 1	ChkStd	38026	7.0942	-7471B	PPB		03 Dec 2018 14:22:25	U	SMPL
PB115239BL - 1	PBS	-316	-0.0451	-7471B	PPB		03 Dec 2018 14:24:44	U	SMPL
PB115239BS - 1	LCSS	17808	3.3296	-7471B	PPB		03 Dec 2018 14:27:02	U	SMPL
J5764-21 - 1	JC-02-113018-A	14604	2.7330	-7471B	PPB		03 Dec 2018 14:29:19	U	SMPL
J5764-23 - 1	JC-02-113018-B	75263	14.0278	-7471B	PPB		03 Dec 2018 14:31:37	U	SMPL
J5764-25 - 1	JC-02-113018-C	15789	2.9537	-7471B	PPB		03 Dec 2018 14:33:55	U	SMPL
J5764-27 - 1	JC-02-113018-D	5573	1.0514	-7471B	PPB		03 Dec 2018 14:36:15	U	SMPL
J5764-29 - 1	JC-02-113018-E	3106	0.5921	-7471B	PPB		03 Dec 2018 14:38:31	U	SMPL
CCV54 - 1	CCV54	26758	4.9961	-7471B	PPB		03 Dec 2018 14:40:47	U	SMPL
CCB54 - 1	CCB54	-220	-0.0272	-7471B	PPB		03 Dec 2018 14:43:03	U	SMPL
J5775-03 - 1	TR-04-113018	1229	0.2426	-7471B	PPB		03 Dec 2018 14:45:20	U	SMPL
J6148-01 - 1	TP-MH-F	42121	7.8567	-7471B	PPB		03 Dec 2018 14:47:36	U	SMPL
J6148-03 - 1	TP-MH-J	31452	5.8701	-7471B	PPB		03 Dec 2018 14:49:51	U	SMPL
J6148-05 - 1	TP-MH-J	31714	5.9189	-7471B	PPB		03 Dec 2018 14:52:10	U	SMPL
J6148-06 - 1	TP-MH-F	48445	9.0342	-7471B	PPB		03 Dec 2018 14:54:28	U	SMPL
J6148-06DUP - 1	TP-MH-FDUP	48602	9.0635	-7471B	PPB		03 Dec 2018 14:56:47	U	SMPL
J6148-06MS - 1	TP-MH-FMS	73182	13.6403	-7471B	PPB		03 Dec 2018 15:01:52	U	SMPL
J6148-06MSD - 1	TP-MH-FMSD	74456	13.8775	-7471B	PPB		03 Dec 2018 15:04:10	U	SMPL
J6160-01 - 1	TEST-PIT	-257	-0.0341	-7471B	PPB		03 Dec 2018 15:06:30	U	SMPL
J5764-23DLX2 - 1		38213	7.1290	-7471B	PPB		03 Dec 2018 15:08:48	U	SMPL
CCV55 - 1	CCV55	27049	5.0503	-7471B	PPB		03 Dec 2018 15:11:05	U	SMPL
CCB55 - 1	CCB55	-242	-0.0313	-7471B	PPB		03 Dec 2018 15:13:23	U	SMPL
J6148-06LX5 - 1		9268	1.7394	-7471B	PPB		03 Dec 2018 15:15:40	U	SMPL
J6148-06A - 1		68692	12.8042	-7471B	PPB		03 Dec 2018 15:22:38	U	SMPL
J6148-06DLX2 - 1		24171	4.5144	-7471B	PPB		03 Dec 2018 15:24:54	U	SMPL
J6148-06DUPDLX2 - 1		24652	4.6040	-7471B	PPB		03 Dec 2018 15:27:13	U	SMPL
J6148-06MSDLX2 - 1		36254	6.7643	-7471B	PPB		03 Dec 2018 15:32:19	U	SMPL
J6148-06MSDDLX2 - 1		36287	6.7704	-7471B	PPB		03 Dec 2018 15:34:35	U	SMPL
J6148-06LDLX10 - 1		4890	0.9243	-7471B	PPB		03 Dec 2018 15:36:54	U	SMPL
J6148-06ADLX2 - 1		35049	6.5399	-7471B	PPB		03 Dec 2018 15:39:13	U	SMPL
CCV56 - 1	CCV56	27076	5.0553	-7471B	PPB		03 Dec 2018 15:41:31	U	SMPL
CCB56 - 1	CCB56	-221	-0.0274	-7471B	PPB		03 Dec 2018 15:43:49	U	SMPL