



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99542

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	81.2	12/03/2018	11:25
2	J6150-01		1	7	>100.0	>212.0	>212.0	12/03/2018	11:50
3	J6150-01DUP		1	7	>100.0	>212.0	>212.0	12/03/2018	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB99542

WorkList Name : fp12318

WorkList ID : 119840

Date : 12/3/2018 10:57:25 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/05/2018	Solid	J6150-01	C	Flash Point	Cool 4 deg C	PSEG01	L31	WASTE-OIL-TOTES	11/30/2018
									1010A

Date/Time 12/3/18 11:00 AM
 Received by: KS
 Relinquished by: CP

Date/Time 12/3/18 12:20 PM
 Received by: CP
 Relinquished by: KS