



Analytical Summary Report

Analysis Method: 1010A
Parameter: Flash Point
Run Number: LB99545

Reviewed By: ketankumar
Supervisor Review By: sweta
Ambient Barometric Pressure (mmHg): 750
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	81.2	12/03/2018	12:45
2	J6153-01		1	7	>100.0	>212.0	>212.0	12/03/2018	13:10
3	J6153-01DUP		1	7	>100.0	>212.0	>212.0	12/03/2018	13:32
4	J6153-02		1	8	>100.0	>212.0	>212.0	12/03/2018	13:50
5	J6153-03		1	6	>100.0	>212.0	>212.0	12/03/2018	14:20
6	J6153-04		1	8	>100.0	>212.0	>212.0	12/03/2018	14:58
7	J6153-05		1	9	>100.0	>212.0	>212.0	12/03/2018	15:30

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

WORKLIST(Hardcopy Internal Chain)

LB 99545

WorkList Name : fp12318-1

WorkList ID : 119841

Date : 12/3/2018 10:58:46 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/13/2018	Water	J6153-01	Flash Point	Cool 4 deg C	TULL02	L22	MH-F	11/29/2018	1010A
12/13/2018	Water	J6153-02	Flash Point	Cool 4 deg C	TULL02	L22	MH-G	11/29/2018	1010A
12/13/2018	Water	J6153-03	Flash Point	Cool 4 deg C	TULL02	L22	MH-J	11/29/2018	1010A
12/13/2018	Water	J6153-04	Flash Point	Cool 4 deg C	TULL02	L22	MH-H	11/29/2018	1010A
12/13/2018	Water	J6153-05	Flash Point	Cool 4 deg C	TULL02	L22	MH-9.41	11/29/2018	1010A

Date/Time 12/3/18 11:00 AM
 Received by: KS
 Relinquished by: CP

Date/Time 12/3/18 5:35
 Received by: CP
 Relinquished by: KS