


LB 99546

ISM02.4

INSTRUMENT ID: CY1

Linear



A= 0.0000e+000

B= 2.3054e-004

C= 2.0642e-002

Rho= 0.9998696

Accept=Accepted

slope
7 = 10183

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.013	0.013	-35	0.000	-35					-
0.2	0.200	0.185	-0.015	713	0.0 %	713					-8
2.5	2.500	2.427	-0.073	10436	0.0 %	10436					-3
5.0	5.000	5.096	0.096	22014	0.0 %	22014					2
7.5	7.500	7.545	0.045	32639	0.0 %	32639					-1
10.0	10.000	9.935	-0.065	43002	0.0 %	43002					-1

NS
12/3/18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-35	-		0 ISM02.4	PPB	12/3/2018 16:09	Std
	0.2 S0.2	713	-		0.2 ISM02.4	PPB	12/3/2018 16:12	Std
	2.5 S2.5	10436	-		2.5 ISM02.4	PPB	12/3/2018 16:14	Std
	5 S5	22014	-		5 ISM02.4	PPB	12/3/2018 16:16	Std
	7.5 S7.5	32639	-		7.5 ISM02.4	PPB	12/3/2018 16:18	Std
	10 S10	43002	-		10 ISM02.4	PPB	12/3/2018 16:23	Std
ICV	ICV	16889	3.9143 -		ISM02.4	PPB	12/3/2018 16:26	SMPL
ICB	ICB	-123	-0.0077 -		ISM02.4	PPB	12/3/2018 16:28	SMPL
CCV057	CCV057	21830	5.0534 -		ISM02.4	PPB	12/3/2018 16:31	SMPL
CCB057	CCB057	-122	-0.0075 -		ISM02.4	PPB	12/3/2018 16:33	SMPL
PB115258BL	PBW258	-65	0.0057 -		ISM02.4	PPB	12/3/2018 16:35	SMPL
J6138-07	MX2F14	21924	5.0751 -		ISM02.4	PPB	12/3/2018 16:40	SMPL
J6138-08	MX2F14D	22370	5.1779 -		ISM02.4	PPB	12/3/2018 16:42	SMPL
J6138-09	MX2F14S	26477	6.1248 -		ISM02.4	PPB	12/3/2018 16:44	SMPL
CCV058	CCV058	21463	4.9688 -		ISM02.4	PPB	12/3/2018 16:47	SMPL
CCB058	CCB058	-132	-0.0098 -		ISM02.4	PPB	12/3/2018 16:49	SMPL