

LB99548


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47682



A= 0.0000e+000

B= 2.0952e-004

C= 4.6849e-003

Rho= 0.9999632

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.008	0.008	16	0.000	16					
0.2	0.200	0.165	-0.035	766	0.0 %	766					-18
2.5	2.500	2.508	0.008	11949	0.0 %	11949					0
5.0	5.000	5.059	0.059	24121	0.0 %	24121					1
7.5	7.500	7.465	-0.035	35607	0.0 %	35607					0
10.0	10.000	9.995	-0.005	47682	0.0 %	47682					0

ns
12/3/18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	16	-		0 ISM02.4	PPB	12/3/2018 18:52	Std
	0.2 S0.2	766	-		0.2 ISM02.4	PPB	12/3/2018 18:54	Std
	2.5 S2.5	11949	-		2.5 ISM02.4	PPB	12/3/2018 18:56	Std
	5 S5	24121	-		5 ISM02.4	PPB	12/3/2018 18:59	Std
	7.5 S7.5	35607	-		7.5 ISM02.4	PPB	12/3/2018 19:01	Std
	10 S10.0	47682	-		10 ISM02.4	PPB	12/3/2018 19:03	Std
ICV	ICV	18701	3.9229 -		ISM02.4	PPB	12/3/2018 19:06	SMPL
ICB	ICB	-57	-0.0073 -		ISM02.4	PPB	12/3/2018 19:08	SMPL
CCV059	CCV059	23359	4.8989 -		ISM02.4	PPB	12/3/2018 19:11	SMPL
CCB059	CCB059	-73	-0.0106 -		ISM02.4	PPB	12/3/2018 19:13	SMPL
PB115252BL	PBS252	38	0.0126 -		ISM02.4	PPB	12/3/2018 19:15	SMPL
J6138-19	MX2F18	18964	3.978 -		ISM02.4	PPB	12/3/2018 19:17	SMPL
J6138-20	MX2F18D	19447	4.0792 -		ISM02.4	PPB	12/3/2018 19:22	SMPL
J6138-21	MX2F18S	30511	6.3973 -		ISM02.4	PPB	12/3/2018 19:24	SMPL
CCV060	CCV060	23131	4.8511 -		ISM02.4	PPB	12/3/2018 19:26	SMPL
CCB060	CCB060	-85	-0.0131 -		ISM02.4	PPB	12/3/2018 19:29	SMPL