



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99560

Slope : 98.5

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.5	4.01	12/04/2018	11:55
2	CAL2	1	Water			20.3	7.00	12/04/2018	11:57
3	CAL3	1	Water			20.4	10.04	12/04/2018	11:59
4	ICV	1	Water			20.2	7.01	12/04/2018	12:01
5	CCV1	1	Water			20.2	2.00	12/04/2018	12:03
6	J6207-01	1	Water			21.9	7.04	12/04/2018	12:05
7	J6207-01DUP	1	Water			22.0	7.05	12/04/2018	12:07
8	J6209-01	1	Water			10.6	6.81	12/04/2018	12:09
9	J6209-03	1	Water			11.9	6.90	12/04/2018	12:11
10	J6209-05	1	Water			13.1	6.96	12/04/2018	12:13
11	J6209-07	1	Water			14.4	7.01	12/04/2018	12:15
12	CCV2	1	Water			20.3	12.00	12/04/2018	12:17

WORKLIST(Hardcopy Internal Chain)

LB 99560

WorkList Name : pH120418

WorkList ID : 119908

Date : 12/4/2018 11:39:58 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/30/2018	Water	J6207-01	pH	Cool 4 deg C	MINO01	L22	OUTFALL	11/30/2018	9040C
12/03/2018	Water	J6209-01	pH	Cool 4 deg C	TULL02	L22	12.68	12/03/2018	9040C
12/03/2018	Water	J6209-03	pH	Cool 4 deg C	TULL02	L22	12.65	12/03/2018	9040C
12/03/2018	Water	J6209-05	pH	Cool 4 deg C	TULL02	L22	HMF-NORTH	12/03/2018	9040C
12/03/2018	Water	J6209-07	pH	Cool 4 deg C	TULL02	L22	HMF-SOUTH	12/03/2018	9040C

Date/Time 12/04/18 11:45
 Received by: SP
 Relinquished by: CP

Date/Time 12/04/18 12:20
 Received by: CP
 Relinquished by: SP