

LB99573


7471B

INSTRUMENT ID : LV1

Linear

 μ Abs.:

53457



A= 0.0000e+000

B= 1.8615e-004

C= 6.6909e-002

Rho= 0.9998922

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.069	0.069	11	0.000	11				
0.2	0.200	0.217	0.017	808	0.0 %	808				
2.5	2.500	2.393	-0.107	12495	0.0 %	12495				
5.0	5.000	4.982	-0.018	26403	0.0 %	26403				
7.5	7.500	7.521	0.021	40042	0.0 %	40042				
10.0	10.000	10.018	0.018	53457	0.0 %	53457				

LB99573 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Dec 2018 15:30:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.2	11	-	0.0000	7471B	PPB	04 Dec 2018 15:34:59	S	Std
0.2 - 1	52.5	808	-	0.2000	7471B	PPB	04 Dec 2018 15:37:17	S	Std
2.5 - 1	55	12495	-	2.5000	7471B	PPB	04 Dec 2018 15:39:34	S	Std
5.0 - 1	57.5	26403	-	5.0000	7471B	PPB	04 Dec 2018 15:41:53	S	Std
7.5 - 1	57.5	40042	-	7.5000	7471B	PPB	04 Dec 2018 15:44:10	S	Std
10.0 - 1	57.5	53457	-	10.0000	7471B	PPB	04 Dec 2018 15:46:27	S	Std
ICV04 - 1	ICV04	21537	4.0761	-	7471B	PPB	04 Dec 2018 15:49:13	U	SMPL
ICB04 - 1	ICB04	-207	0.0284	-	7471B	PPB	04 Dec 2018 15:51:29	U	SMPL
CCV68 - 1	CCV68	27700	5.2233	-	7471B	PPB	04 Dec 2018 15:53:45	U	SMPL
CCB68 - 1	CCB68	-65	0.0548	-	7471B	PPB	04 Dec 2018 15:59:10	U	SMPL
CRA - 1	CRA	800	0.2158	-	7471B	PPB	04 Dec 2018 16:01:25	U	SMPL
HighStd - 1	HighStd	52834	9.9021	-	7471B	PPB	04 Dec 2018 16:03:41	U	SMPL
ChkStd - 1	ChkStd	35194	6.6184	-	7471B	PPB	04 Dec 2018 16:05:57	U	SMPL
PB115296BL - 1	PBS	-254	0.0196	-	7471B	PPB	04 Dec 2018 16:08:16	U	SMPL
PB115296BS - 1	LCSS	19072	3.6172	-	7471B	PPB	04 Dec 2018 16:10:34	U	SMPL
J6168-03 - 1	S1	11191	2.1501	-	7471B	PPB	04 Dec 2018 16:12:51	U	SMPL
J6168-08 - 1	S2	9926	1.9147	-	7471B	PPB	04 Dec 2018 16:15:09	U	SMPL
J6168-14 - 1	S3	11196	2.1511	-	7471B	PPB	04 Dec 2018 16:17:27	U	SMPL
J6168-19 - 1	S4	2322	0.4992	-	7471B	PPB	04 Dec 2018 16:19:45	U	SMPL
J6168-23 - 1	S5	20335	3.8523	-	7471B	PPB	04 Dec 2018 16:22:01	U	SMPL
CCV69 - 1	CCV69	27503	5.1867	-	7471B	PPB	04 Dec 2018 16:24:17	U	SMPL
CCB69 - 1	CCB69	-200	0.0297	-	7471B	PPB	04 Dec 2018 16:26:33	U	SMPL
J6168-28 - 1	S6	5903	1.1658	-	7471B	PPB	04 Dec 2018 16:28:50	U	SMPL
J6184-01 - 1	RO-331	7828	1.5241	-	7471B	PPB	04 Dec 2018 16:31:06	U	SMPL
J6192-01 - 1	HD-02-120318	3107	0.6453	-	7471B	PPB	04 Dec 2018 16:33:23	U	SMPL
J6198-01 - 1	SA-06-120318	2577	0.5466	-	7471B	PPB	04 Dec 2018 16:35:39	U	SMPL
J6201-01 - 1	72-11932-72-11967	3279	0.6773	-	7471B	PPB	04 Dec 2018 16:37:56	U	SMPL
J6204-02 - 1	HLSP-SB-1A-(0-2)	5219	1.0384	-	7471B	PPB	04 Dec 2018 16:40:13	U	SMPL
J6204-04 - 1	HLSP-SB-1B-(2-6)	1373	0.3225	-	7471B	PPB	04 Dec 2018 16:42:31	U	SMPL
J6204-06 - 1	HLSP-SB-2A-(0-2)	8052	1.5658	-	7471B	PPB	04 Dec 2018 16:44:48	U	SMPL
J6204-08 - 1	HLSP-SB-2B-(2-9.75)	273	0.1177	-	7471B	PPB	04 Dec 2018 16:47:07	U	SMPL
J6204-11 - 1	HLSP-SB-7A-(0-2)	21496	4.0685	-	7471B	PPB	04 Dec 2018 16:49:23	U	SMPL
CCV70 - 1	CCV70	27454	5.1776	-	7471B	PPB	04 Dec 2018 16:51:39	U	SMPL
CCB70 - 1	CCB70	-146	0.0397	-	7471B	PPB	04 Dec 2018 16:56:44	U	SMPL
J6204-13 - 1	HLSP-SB-7B-(2-6)	14708	2.8048	-	7471B	PPB	04 Dec 2018 16:59:00	U	SMPL
J6204-15 - 1	HLSP-SB-8A-(0-2)	4783	0.9573	-	7471B	PPB	04 Dec 2018 17:01:16	U	SMPL
J6204-17 - 1	HLSP-SB-8B-(2-9.75)	6602	1.2959	-	7471B	PPB	04 Dec 2018 17:03:32	U	SMPL
J6204-19 - 1	HLSP-DUP01-112918	5049	1.0068	-	7471B	PPB	04 Dec 2018 17:16:56	U	SMPL
J6204-19DUP - 1	HLSP-DUP01-112918DUP	5073	1.0113	-	7471B	PPB	04 Dec 2018 17:19:12	U	SMPL
J6204-19MS - 1	HLSP-DUP01-112918MS	23840	4.5048	-	7471B	PPB	04 Dec 2018 17:21:29	U	SMPL
J6204-19MSD - 1	HLSP-DUP01-112918MSD	24094	4.5521	-	7471B	PPB	04 Dec 2018 17:23:47	U	SMPL
PB115297BL - 1	PBS	-253	0.0198	-	7471B	PPB	04 Dec 2018 17:26:06	U	SMPL
PB115297BS - 1	LCSS	18755	3.5582	-	7471B	PPB	04 Dec 2018 17:28:25	U	SMPL
J6206-01 - 1	HLSP-SB-5A-(0-2)	2858	0.5989	-	7471B	PPB	04 Dec 2018 17:30:41	U	SMPL
CCV71 - 1	CCV71	27369	5.1617	-	7471B	PPB	04 Dec 2018 17:32:58	U	SMPL
CCB71 - 1	CCB71	-256	0.0193	-	7471B	PPB	04 Dec 2018 17:35:14	U	SMPL
J6206-03 - 1	HLSP-SB-5B-(2-8.75)	1034	0.2594	-	7471B	PPB	04 Dec 2018 17:37:32	U	SMPL
J6206-03DUP - 1	HLSP-SB-5B-(2-8.75)DUP	1025	0.2577	-	7471B	PPB	04 Dec 2018 17:39:48	U	SMPL
J6206-05 - 1	J6206-03MS	20126	3.8134	-	7471B	PPB	04 Dec 2018 17:42:05	U	SMPL
J6206-07 - 1	J6206-03MSD	19047	3.6126	-	7471B	PPB	04 Dec 2018 17:44:22	U	SMPL
J6206-09 - 1	HLSP-SB-6A-(0-2)	1782	0.3986	-	7471B	PPB	04 Dec 2018 17:46:38	U	SMPL
J6206-11 - 1	HLSP-SB-6B-(2-5)	140	0.0930	-	7471B	PPB	04 Dec 2018 17:48:56	U	SMPL
J6206-13 - 1	HLSP-SB-4A-(0-2)	15604	2.9716	-	7471B	PPB	04 Dec 2018 17:51:14	U	SMPL
J6206-15 - 1	HLSP-SB-4B-(2-6)	745	0.2056	-	7471B	PPB	04 Dec 2018 17:53:32	U	SMPL
J6206-17 - 1	HLSP-SB-3A-(0-2)	1828	0.4072	-	7471B	PPB	04 Dec 2018 17:55:50	U	SMPL
J6206-19 - 1	HLSP-SB-3B-(2-7.5)	366	0.1350	-	7471B	PPB	04 Dec 2018 17:58:06	U	SMPL
CCV72 - 1	CCV72	27699	5.2232	-	7471B	PPB	04 Dec 2018 18:00:23	U	SMPL
CCB72 - 1	CCB72	94	0.0844	-	7471B	PPB	04 Dec 2018 18:02:40	U	SMPL
J6204-19LX5 - 1		564	0.1719	-	7471B	PPB	04 Dec 2018 18:04:58	U	SMPL
J6204-19A - 1		26785	5.0530	-	7471B	PPB	04 Dec 2018 18:07:14	U	SMPL
J6206-03LX5 - 1		-72	0.0535	-	7471B	PPB	04 Dec 2018 18:09:30	U	SMPL
J6206-03A - 1		23198	4.3853	-	7471B	PPB	04 Dec 2018 18:11:48	U	SMPL
CCV73 - 1	CCV73	26613	5.0210	-	7471B	PPB	04 Dec 2018 18:14:05	U	SMPL
CCB73 - 1	CCB73	-227	0.0247	-	7471B	PPB	04 Dec 2018 18:16:22	U	SMPL