

LB99576


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

34702



A= 0.0000e+000

B= 2.8717e-004

C= -9.2216e-002

Rho= 0.9997992

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.043	-0.043	173	0.000	173				
0.05	0.050	0.048	-0.002	490	0.0 %	490				
0.2	0.200	0.186	-0.014	970	0.0 %	970				
2.5	2.500	2.483	-0.017	8967	0.0 %	8967				
5.0	5.000	5.095	0.095	18063	0.0 %	18063				
7.5	7.500	7.607	0.107	26809	0.0 %	26809				
10.0	10.000	9.873	-0.127	34702	0.0 %	34702				

LB99576 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 04 Dec 2018 19:25:04

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	173	-	0.0000245.1		PPB	04 Dec 2018 19:31:48	S	Std
0.05 - 1	50.05	490	-	0.0500245.1		PPB	04 Dec 2018 19:34:06	S	Std
0.2 - 1	50.2	970	-	0.2000245.1		PPB	04 Dec 2018 19:36:24	S	Std
2.5 - 1	52.5	8967	-	2.5000245.1		PPB	04 Dec 2018 19:38:42	S	Std
5.0 - 1	55	18063	-	5.0000245.1		PPB	04 Dec 2018 19:40:59	S	Std
7.5 - 1	57.5	26809	-	7.5000245.1		PPB	04 Dec 2018 19:43:16	S	Std
10.0 - 1	510	34702	-	10.0000245.1		PPB	04 Dec 2018 19:48:13	S	Std
ICV05 - 1	ICV05	14183	3.9808	-245.1		PPB	04 Dec 2018 19:51:04	U	SMPL
ICB05 - 1	ICB05	369	0.0138	-245.1		PPB	04 Dec 2018 19:53:20	U	SMPL
CCV74 - 1	CCV74	18375	5.1846	-245.1		PPB	04 Dec 2018 19:55:36	U	SMPL
CCB74 - 1	CCB74	366	0.0129	-245.1		PPB	04 Dec 2018 19:57:52	U	SMPL
CRA - 1	CRA	506	0.0531	-245.1		PPB	04 Dec 2018 20:00:09	U	SMPL
HighStd - 1	HighStd	35065	9.9775	-245.1		PPB	04 Dec 2018 20:02:25	U	SMPL
ChkStd - 1	ChkStd	24754	7.0165	-245.1		PPB	04 Dec 2018 20:04:41	U	SMPL
PB115336BL - 1	PBW	366	0.0129	-245.1		PPB	04 Dec 2018 20:09:35	U	SMPL
PB115336BS - 1	LCSW	13836	3.8811	-245.1		PPB	04 Dec 2018 20:11:52	U	SMPL
J6128-03 - 1	112718-DBRI-E-U-B	-111	-0.1241	-245.1		PPB	04 Dec 2018 20:16:50	U	SMPL
J6128-04 - 1	112718-DBRI-E-U-S	-106	-0.1227	-245.1		PPB	04 Dec 2018 20:19:07	U	SMPL
J6128-05 - 1	112718-TIPC-F-D-B	-108	-0.1232	-245.1		PPB	04 Dec 2018 20:21:24	U	SMPL
J6128-06 - 1	112718-TIPC-F-D-S	-105	-0.1224	-245.1		PPB	04 Dec 2018 20:23:42	U	SMPL
J6128-07 - 1	112718-TIPC-F-U-B	-110	-0.1238	-245.1		PPB	04 Dec 2018 20:25:58	U	SMPL
CCV75 - 1	CCV75	18459	5.2087	-245.1		PPB	04 Dec 2018 20:28:14	U	SMPL
CCB75 - 1	CCB75	367	0.0132	-245.1		PPB	04 Dec 2018 20:30:30	U	SMPL
J6128-08 - 1	112718-TIPC-F-U-S	-132	-0.1301	-245.1		PPB	04 Dec 2018 20:32:47	U	SMPL
J6129-01 - 1	112818-TIPC-F-D-B	-129	-0.1293	-245.1		PPB	04 Dec 2018 20:35:03	U	SMPL
J6129-08 - 1	112818-DBRI-F-U-S	-131	-0.1298	-245.1		PPB	04 Dec 2018 20:37:19	U	SMPL
J6129-09 - 1	112818-CHDE-F-D-B	-136	-0.1313	-245.1		PPB	04 Dec 2018 20:39:36	U	SMPL
J6129-10 - 1	112818-CHDE-F-D-M	-124	-0.1278	-245.1		PPB	04 Dec 2018 20:41:53	U	SMPL
J6129-11 - 1	112818-CHDE-F-D-S	-119	-0.1264	-245.1		PPB	04 Dec 2018 20:44:10	U	SMPL
J6129-12 - 1	112818-CHDE-F-U-B	-119	-0.1264	-245.1		PPB	04 Dec 2018 20:46:27	U	SMPL
J6129-13 - 1	112818-CHDE-F-U-M	270	-0.0147	-245.1		PPB	04 Dec 2018 20:48:45	U	SMPL
PB115337BL - 1	PBW	365	0.0126	-245.1		PPB	04 Dec 2018 20:53:37	U	SMPL
PB115337BS - 1	LCSW	13554	3.8001	-245.1		PPB	04 Dec 2018 20:55:53	U	SMPL
CCV76 - 1	CCV76	18676	5.2710	-245.1		PPB	04 Dec 2018 20:58:09	U	SMPL
CCB76 - 1	CCB76	348	0.0077	-245.1		PPB	04 Dec 2018 21:00:26	U	SMPL
J6135-01 - 1	20180972-MANASQUAN-H20	-104	-0.1221	-245.1		PPB	04 Dec 2018 21:02:43	U	SMPL
J6135-01DUP - 1	20180972-MANASQUAN-H20DUP	-123	-0.1275	-245.1		PPB	04 Dec 2018 21:04:59	U	SMPL
J6135-01MS - 1	20180972-MANASQUAN-H20MS	15065	4.2340	-245.1		PPB	04 Dec 2018 21:07:15	U	SMPL
J6135-01MSD - 1	20180972-MANASQUAN-H20MSD	15356	4.3176	-245.1		PPB	04 Dec 2018 21:09:32	U	SMPL
J6151-01 - 1	112918-DBRI-E-D-M	-134	-0.1307	-245.1		PPB	04 Dec 2018 21:14:29	U	SMPL
J6151-02 - 1	112918-DBRI-E-U-M	-139	-0.1321	-245.1		PPB	04 Dec 2018 21:16:46	U	SMPL
J6151-03 - 1	112918-TIPC-E-D-M	-143	-0.1333	-245.1		PPB	04 Dec 2018 21:19:04	U	SMPL
J6151-04 - 1	112918-TIPC-E-U-M	-135	-0.1310	-245.1		PPB	04 Dec 2018 21:21:21	U	SMPL
J6213-01 - 1	113018-JETT-E-D-B	-122	-0.1273	-245.1		PPB	04 Dec 2018 21:23:39	U	SMPL
J6213-02 - 1	113018-JETT-E-D-M	-117	-0.1258	-245.1		PPB	04 Dec 2018 21:25:55	U	SMPL
CCV77 - 1	CCV77	18436	5.2021	-245.1		PPB	04 Dec 2018 21:28:12	U	SMPL
CCB77 - 1	CCB77	369	0.0138	-245.1		PPB	04 Dec 2018 21:30:28	U	SMPL
J6213-03 - 1	113018-JETT-E-D-S	-144	-0.1336	-245.1		PPB	04 Dec 2018 21:32:46	U	SMPL
J6213-04 - 1	113018-JETT-E-U-B	-127	-0.1287	-245.1		PPB	04 Dec 2018 21:35:02	U	SMPL
J6213-05 - 1	113018-JETT-E-U-M	-135	-0.1310	-245.1		PPB	04 Dec 2018 21:37:19	U	SMPL
J6213-06 - 1	113018-JETT-E-U-S	-117	-0.1258	-245.1		PPB	04 Dec 2018 21:39:35	U	SMPL
J6213-07 - 1	113018-TIPC-E-D-M	-135	-0.1310	-245.1		PPB	04 Dec 2018 21:41:52	U	SMPL
J6213-08 - 1	113018-TIPC-E-U-M	-124	-0.1278	-245.1		PPB	04 Dec 2018 21:44:09	U	SMPL
J6135-01LX5 - 1		-53	-0.1074	-245.1		PPB	04 Dec 2018 21:46:27	U	SMPL
J6135-01A - 1		17272	4.8678	-245.1		PPB	04 Dec 2018 21:48:45	U	SMPL
CCV78 - 1	CCV78	18917	5.3402	-245.1		PPB	04 Dec 2018 21:51:03	U	SMPL
CCB78 - 1	CCB78	-245	-0.1626	-245.1		PPB	04 Dec 2018 21:53:20	U	SMPL