



Analytical Summary Report

Analysis Method: 9045D

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB99585

Slope : 98.5

BalanceID: _____

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	12/04/2018	17:30
2	CAL2	1	Water			20.3	7.00	12/04/2018	17:32
3	CAL3	1	Water			20.1	10.04	12/04/2018	17:34
4	ICV	1	Water			20.1	7.01	12/04/2018	17:36
5	CCV1	1	Water			20.2	2.00	12/04/2018	17:38
6	J6182-18	1	Solid	20.02	20	20.0	11.44	12/04/2018	17:40
7	J6182-18DUP	1	Solid	20.03	20	20.1	11.45	12/04/2018	17:42
8	J6188-02	1	Solid	20.03	20	18.9	7.32	12/04/2018	17:44
9	J6194-02	1	Solid	20.04	20	20.0	7.71	12/04/2018	17:46
10	J6194-04	1	Solid	20.02	20	19.8	8.96	12/04/2018	17:48
11	J6202-02	1	Solid	20.02	20	20.3	8.12	12/04/2018	17:50
12	J6208-02	1	Solid	20.03	20	21.1	6.93	12/04/2018	17:52
13	J6208-06	1	Solid	20.02	20	21.2	8.10	12/04/2018	17:54
14	J6208-10	1	Solid	20.02	20	21.3	7.91	12/04/2018	17:56
15	J6208-14	1	Solid	20.04	20	21.2	8.13	12/04/2018	17:58
16	CCV2	1	Water			20.1	12.00	12/04/2018	18:00
17	J6208-18	1	Solid	20.02	20	21.1	8.45	12/04/2018	18:02
18	J6208-22	1	Solid	20.03	20	21.5	8.24	12/04/2018	18:04
19	J6210-08	1	Solid	20.03	20	20.2	8.35	12/04/2018	18:06
20	J6210-08DUP	1	Solid	20.04	20	20.3	8.36	12/04/2018	18:08

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	J6210-10	1	Solid	20.02	20	20.0	6.40	12/04/2018	18:10
22	J6214-02	1	Solid	20.02	20	20.5	7.51	12/04/2018	18:12
23	J6214-14	1	Solid	20.03	20	20.7	6.44	12/04/2018	18:14
24	J6219-04	1	Solid	20.02	20	19.7	6.83	12/04/2018	18:16
25	J6233-02	1	Solid	20.04	20	19.5	6.76	12/04/2018	18:18
26	J6233-04	1	Solid	20.02	20	20.3	5.08	12/04/2018	18:20
27	CCV3	1	Water			20.2	2.00	12/04/2018	18:22

WORKLIST(Hardcopy Internal Chain)

LB 99585

WorkList Name : CORROSIONITY120418

WorkList ID : 119907

Date : 12/4/2018 11:39:51 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/29/2018	Solid	J6182-18	Corrosivity	Cool 4 deg C	ENVI77	L21	OXS01-02	11/29/2018	9045D
12/04/2018	Solid	J6188-02	Corrosivity	Cool 4 deg C	PSEG05	L31	BU-03-120418-A	12/04/2018	9045D
12/04/2018	Solid	J6194-02	Corrosivity	Cool 4 deg C	PSEG05	L31	OK-02-120418-A	12/04/2018	9045D
12/04/2018	Solid	J6194-04	Corrosivity	Cool 4 deg C	PSEG05	L31	OK-02-120418-B	12/04/2018	9045D
12/04/2018	Solid	J6202-02	Corrosivity	Cool 4 deg C	PSEG01	L31	SAMPLE-1	12/04/2018	9045D
12/03/2018	Solid	J6208-02	Corrosivity	Cool 4 deg C	TULL02	L61	TP-1	12/03/2018	9045D
12/03/2018	Solid	J6208-06	Corrosivity	Cool 4 deg C	TULL02	L61	TP-2	12/03/2018	9045D
12/03/2018	Solid	J6208-10	Corrosivity	Cool 4 deg C	TULL02	L61	TP-3	12/03/2018	9045D
12/03/2018	Solid	J6208-14	Corrosivity	Cool 4 deg C	TULL02	L61	TP-4	12/03/2018	9045D
12/03/2018	Solid	J6208-18	Corrosivity	Cool 4 deg C	TULL02	L61	TP-5	12/03/2018	9045D
12/03/2018	Solid	J6208-22	Corrosivity	Cool 4 deg C	TULL02	L61	TP-6	12/03/2018	9045D
11/29/2018	Solid	J6210-08	Corrosivity	Cool 4 deg C	BRIN05	L22	C-1-112918	11/29/2018	9045D
11/29/2018	Solid	J6210-10	Corrosivity	Cool 4 deg C	BRIN05	L22	C-2-112918	11/29/2018	9045D
12/03/2018	Solid	J6214-02	Corrosivity	Cool 4 deg C	AKRF02	L31	SB-8-(13-15)	12/03/2018	9045D
12/03/2018	Solid	J6214-14	Corrosivity	Cool 4 deg C	AKRF02	L31	SB-1-(0-2)	12/03/2018	9045D
12/04/2018	Solid	J6219-04	Corrosivity	Cool 4 deg C	PSEG01	L31	NM-342-135-120418	12/04/2018	9045D
12/04/2018	Solid	J6233-02	Corrosivity	Cool 4 deg C	PSEG01	L31	SB-1	12/04/2018	9045D
12/04/2018	Solid	J6233-04	Corrosivity	Cool 4 deg C	PSEG01	L31	SB-2	12/04/2018	9045D

Date/Time 12/04/18 5:20

Received by: SP

Relinquished by: CP

Date/Time

Received by:

Relinquished by:

12/04/18 6:30

CP