


7470A

LB99587
INSTRUMENT ID: CV1

Linear

 μ Abs.:

38609



A= 0.0000e+000

B= 2.6516e-004

C= -1.7033e-001

Rho= 0.9998356

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	588	0.000	588				
0.2	0.200	0.192	-0.008	1367	0.0 %	1367				
2.5	2.500	2.555	0.055	10278	0.0 %	10278				
5.0	5.000	5.034	0.034	19628	0.0 %	19628				
7.5	7.500	7.366	-0.134	28423	0.0 %	28423				
10.0	10.000	10.067	0.067	38609	0.0 %	38609				

LB99587 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Dec 2018 10:50:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	588	-	0.00007470A	PPB		05 Dec 2018 10:54:48	S	Std
0.2 - 1	50.2	1367	-	0.20007470A	PPB		05 Dec 2018 10:56:57	S	Std
2.5 - 1	52.5	10278	-	2.50007470A	PPB		05 Dec 2018 10:59:05	S	Std
5.0 - 1	55	19628	-	5.00007470A	PPB		05 Dec 2018 11:01:14	S	Std
7.5 - 1	57.5	28423	-	7.50007470A	PPB		05 Dec 2018 11:03:22	S	Std
10.0 - 1	510	38609	-	10.00007470A	PPB		05 Dec 2018 11:09:12	S	Std
ICV06 - 1	ICV06	14731	3.7357	-7470A	PPB		05 Dec 2018 11:11:53	U	SMPL
ICB06 - 1	ICB06	486	-0.0415	-7470A	PPB		05 Dec 2018 11:14:00	U	SMPL
CCV79 - 1	CCV79	19692	5.0511	-7470A	PPB		05 Dec 2018 11:16:07	U	SMPL
CCB79 - 1	CCB79	487	-0.0412	-7470A	PPB		05 Dec 2018 11:18:14	U	SMPL
CRA - 1	CRA	1346	0.1866	-7470A	PPB		05 Dec 2018 11:20:21	U	SMPL
HighStd - 1	HighStd	38848	10.1304	-7470A	PPB		05 Dec 2018 11:22:28	U	SMPL
ChkStd - 1	ChkStd	26375	6.8231	-7470A	PPB		05 Dec 2018 11:24:35	U	SMPL
PB115338BL - 1	PBW	443	-0.0529	-7470A	PPB		05 Dec 2018 11:26:43	U	SMPL
PB115338BS - 1	LCSW	14411	3.6508	-7470A	PPB		05 Dec 2018 11:28:51	U	SMPL
J6168-04 - 1	S1	525	-0.0311	-7470A	PPB		05 Dec 2018 11:30:59	U	SMPL
J6168-09 - 1	S2	567	-0.0200	-7470A	PPB		05 Dec 2018 11:33:07	U	SMPL
J6168-15 - 1	S3	584	-0.0155	-7470A	PPB		05 Dec 2018 11:35:15	U	SMPL
J6168-20 - 1	S4	591	-0.0136	-7470A	PPB		05 Dec 2018 11:37:24	U	SMPL
J6168-24 - 1	S5	546	-0.0256	-7470A	PPB		05 Dec 2018 11:39:31	U	SMPL
CCV80 - 1	CCV80	19693	5.0514	-7470A	PPB		05 Dec 2018 11:41:37	U	SMPL
CCB80 - 1	CCB80	531	-0.0295	-7470A	PPB		05 Dec 2018 11:43:44	U	SMPL
J6168-29 - 1	S6	549	-0.0248	-7470A	PPB		05 Dec 2018 11:45:51	U	SMPL
J6182-18 - 1	OXS01-02	1161	0.1375	-7470A	PPB		05 Dec 2018 11:47:58	U	SMPL
J6184-02 - 1	RO-331	489	-0.0407	-7470A	PPB		05 Dec 2018 11:50:05	U	SMPL
J6184-02DUP - 1	RO-331DUP	571	-0.0189	-7470A	PPB		05 Dec 2018 11:52:12	U	SMPL
J6184-02MS - 1	RO-331MS	13496	3.4082	-7470A	PPB		05 Dec 2018 11:54:20	U	SMPL
J6184-02MSD - 1	RO-331MSD	13069	3.2950	-7470A	PPB		05 Dec 2018 11:56:28	U	SMPL
J6192-02 - 1	HD-02-120318	538	-0.0277	-7470A	PPB		05 Dec 2018 11:58:36	U	SMPL
J6198-02 - 1	SA-06-120318	518	-0.0330	-7470A	PPB		05 Dec 2018 12:00:44	U	SMPL
J6201-02 - 1	72-11932-72-11967	533	-0.0290	-7470A	PPB		05 Dec 2018 12:02:53	U	SMPL
PB115340BL - 1	PBW	627	-0.0041	-7470A	PPB		05 Dec 2018 12:05:00	U	SMPL
CCV81 - 1	CCV81	19325	4.9538	-7470A	PPB		05 Dec 2018 12:07:06	U	SMPL
CCB81 - 1	CCB81	518	-0.0330	-7470A	PPB		05 Dec 2018 12:09:13	U	SMPL
PB115340BS - 1	LCSW	13622	3.4416	-7470A	PPB		05 Dec 2018 12:11:20	U	SMPL
J6204-09 - 1	HLSP-FB1-112918	576	-0.0176	-7470A	PPB		05 Dec 2018 12:13:27	U	SMPL
J6209-01 - 1	12.68	567	-0.0200	-7470A	PPB		05 Dec 2018 12:15:35	U	SMPL
J6209-01DUP - 1	12.68DUP	515	-0.0338	-7470A	PPB		05 Dec 2018 12:17:42	U	SMPL
J6209-01MS - 1	12.68MS	15058	3.8224	-7470A	PPB		05 Dec 2018 12:19:49	U	SMPL
J6209-01MSD - 1	12.68MSD	15058	3.8224	-7470A	PPB		05 Dec 2018 12:21:57	U	SMPL
J6209-03 - 1	12.65	454	-0.0499	-7470A	PPB		05 Dec 2018 12:24:05	U	SMPL
J6209-05 - 1	HMF-NORTH	524	-0.0314	-7470A	PPB		05 Dec 2018 12:26:13	U	SMPL
J6209-07 - 1	HMF-SOUTH	510	-0.0351	-7470A	PPB		05 Dec 2018 12:28:22	U	SMPL
CCV82 - 1	CCV82	19477	4.9941	-7470A	PPB		05 Dec 2018 12:30:29	U	SMPL
CCB82 - 1	CCB82	546	-0.0256	-7470A	PPB		05 Dec 2018 12:32:36	U	SMPL
PB115338TB - 1	PB115292TB	608	-0.0091	-7470A	PPB		05 Dec 2018 12:34:43	U	SMPL
J6184-02LX5 - 1		588	-0.0144	-7470A	PPB		05 Dec 2018 12:36:51	U	SMPL
J6184-02A - 1		15754	4.0069	-7470A	PPB		05 Dec 2018 12:38:58	U	SMPL
J6209-01LX5 - 1		572	-0.0187	-7470A	PPB		05 Dec 2018 12:41:05	U	SMPL
J6209-01A - 1		15829	4.0268	-7470A	PPB		05 Dec 2018 12:43:13	U	SMPL
CCV83 - 1	CCV83	19537	5.0100	-7470A	PPB		05 Dec 2018 12:45:20	U	SMPL
CCB83 - 1	CCB83	515	-0.0338	-7470A	PPB		05 Dec 2018 12:47:28	U	SMPL