



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99590

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	12/05/2018	09:50
2	J6209-01		1	9	>100.0	>212.0	>212.0	12/05/2018	10:15
3	J6209-01DUP		1	9	>100.0	>212.0	>212.0	12/05/2018	10:40
4	J6209-03		1	7	>100.0	>212.0	>212.0	12/05/2018	11:10
5	J6209-05		1	9	>100.0	>212.0	>212.0	12/05/2018	11:35
6	J6209-07		1	8	>100.0	>212.0	>212.0	12/05/2018	11:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB99590

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP12518

WorkList ID : 119939

Date : 12/5/2018 9:13:45 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/17/2018	Water	J6209-01	Flash Point	Cool 4 deg C	TULL02	L22	12.68	12/03/2018	1010A
12/17/2018	Water	J6209-03	Flash Point	Cool 4 deg C	TULL02	L22	12.65	12/03/2018	1010A
12/17/2018	Water	J6209-05	Flash Point	Cool 4 deg C	TULL02	L22	HMF-NORTH	12/03/2018	1010A
12/17/2018	Water	J6209-07	Flash Point	Cool 4 deg C	TULL02	L22	HMF-SOUTH	12/03/2018	1010A

Date/Time 12/5/18 9:20 AM

Received by: KS

Relinquished by: CP

Date/Time 12/5/18 11:55 AM

Received by: CP

Relinquished by: KS