

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB99593

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP69975
calibration std. phosphate 0.5 ppm	WP69974
calibration std. phosphate 0.3 ppm	WP69973
calibration std. 0 ppm	WP69970
phosphate ICV std.	WP69977
phosphate CCV std.	WP69976
calibration std. phosphate 0.1 ppm	WP69972
calibration std. phosphate 0.05 ppm	WP69971
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924
5N sulfuric acid	WP65054
Combined reagent	WP69915
Sodium hydroxide, 1N	WP66924

Intercept: -0.0009Slope: 0.6516Regression: 0.999813

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	12/04/2018	11:34
2	CAL2	0.05	1	50	50	0.033	12/04/2018	11:35
3	CAL3	0.10	1	50	50	0.065	12/04/2018	11:36
4	CAL4	0.30	1	50	50	0.186	12/04/2018	11:36
5	CAL5	0.50	1	50	50	0.331	12/04/2018	11:37
6	CAL6	1.00	1	50	50	0.650	12/04/2018	11:37



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB99593

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.327	0.503	12/04/2018	11:37
2	ICB		1	50	50	0.000	0.001	12/04/2018	11:38
3	CCV1	0.50	1	50	50	0.329	0.506	12/04/2018	11:39
4	CCB1		1	50	50	0.000	0.001	12/04/2018	11:39
5	PB115325BL		1	50	50	0.000	0.001	12/04/2018	11:39
6	PB115325BS	0.50	1	50	50	0.325	0.500	12/04/2018	11:40
7	J6130-01		1	50	50	0.026	0.041	12/04/2018	11:41
8	J6130-01DUP		1	50	50	0.022	0.035	12/04/2018	11:41
9	J6130-01MS	0.50	1	50	50	0.372	0.572	12/04/2018	11:42
10	J6130-01MSD	0.50	1	50	50	0.376	0.578	12/04/2018	11:43
11	J6130-02		1	50	50	0.030	0.047	12/04/2018	11:43
12	J6156-01		1	50	50	1.060	1.628	12/04/2018	11:44
13	J6215-01		1	50	50	0.080	0.124	12/04/2018	11:44
14	J6156-01		5	50	50	0.282	0.434	12/04/2018	11:50
15	CCV2	0.50	1	50	50	0.327	0.503	12/04/2018	11:50
16	CCB2		1	50	50	0.000	0.001	12/04/2018	11:51