


7471B

LB99594
INSTRUMENT ID: CV1

Linear

 μ Abs.:

54705



A= 0.0000e+000

B= 1.8274e-004

C= -9.9980e-003

Rho= 0.9999897

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	-4	0.000	-4				
0.2	0.200	0.210	0.010	1203	0.0 %	1203				
2.5	2.500	2.480	-0.020	13628	0.0 %	13628				
5.0	5.000	5.030	0.030	27581	0.0 %	27581				
7.5	7.500	7.503	0.003	41114	0.0 %	41114				
10.0	10.000	9.987	-0.013	54705	0.0 %	54705				

LB99594 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Dec 2018 14:25:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-4	-	0.0000	7471B	PPB	05 Dec 2018 14:27:21	S	Std
0.2 - 1	50.2	1203	-	0.2000	7471B	PPB	05 Dec 2018 14:29:39	S	Std
2.5 - 1	52.5	13628	-	2.5000	7471B	PPB	05 Dec 2018 14:31:57	S	Std
5.0 - 1	55	27581	-	5.0000	7471B	PPB	05 Dec 2018 14:34:15	S	Std
7.5 - 1	57.5	41114	-	7.5000	7471B	PPB	05 Dec 2018 14:36:32	S	Std
10.0 - 1	510	54705	-	10.0000	7471B	PPB	05 Dec 2018 14:38:50	S	Std
ICV07 - 1	ICV07	20651	3.7638	-	7471B	PPB	05 Dec 2018 14:41:39	U	SMPL
ICB07 - 1	ICB07	-80	-0.0246	-	7471B	PPB	05 Dec 2018 14:43:54	U	SMPL
CCV84 - 1	CCV84	26674	4.8645	-	7471B	PPB	05 Dec 2018 14:46:11	U	SMPL
CCB84 - 1	CCB84	-100	-0.0283	-	7471B	PPB	05 Dec 2018 14:48:27	U	SMPL
CRA - 1	CRA	1157	0.2014	-	7471B	PPB	05 Dec 2018 14:50:45	U	SMPL
HighStd - 1	HighStd	53051	9.6847	-	7471B	PPB	05 Dec 2018 14:53:01	U	SMPL
ChkStd - 1	ChkStd	37458	6.8352	-	7471B	PPB	05 Dec 2018 14:55:17	U	SMPL
PB115341BL - 1	PBS	-69	-0.0226	-	7471B	PPB	05 Dec 2018 14:57:36	U	SMPL
PB115341BS - 1	LCSS	18278	3.3302	-	7471B	PPB	05 Dec 2018 14:59:54	U	SMPL
J6208-01 - 1	TP-1	21023	3.8318	-	7471B	PPB	05 Dec 2018 15:02:11	U	SMPL
J6208-05 - 1	TP-2	26721	4.8731	-	7471B	PPB	05 Dec 2018 15:04:28	U	SMPL
J6208-09 - 1	TP-3	12758	2.3214	-	7471B	PPB	05 Dec 2018 15:06:47	U	SMPL
J6208-13 - 1	TP-4	14686	2.6738	-	7471B	PPB	05 Dec 2018 15:09:06	U	SMPL
J6208-17 - 1	TP-5	33086	6.0362	-	7471B	PPB	05 Dec 2018 15:11:22	U	SMPL
CCV85 - 1	CCV85	26364	4.8078	-	7471B	PPB	05 Dec 2018 15:13:38	U	SMPL
CCB85 - 1	CCB85	-98	-0.0279	-	7471B	PPB	05 Dec 2018 15:15:56	U	SMPL
J6208-21 - 1	TP-6	14032	2.5542	-	7471B	PPB	05 Dec 2018 15:18:13	U	SMPL
J6214-01 - 1	SB-8-(13-15)	9931	1.8048	-	7471B	PPB	05 Dec 2018 15:20:29	U	SMPL
J6214-03 - 1	SB-8-(18-20)	28932	5.2771	-	7471B	PPB	05 Dec 2018 15:22:45	U	SMPL
J6214-05 - 1	SB-9-(13-15)	26089	4.7576	-	7471B	PPB	05 Dec 2018 15:25:02	U	SMPL
J6214-07 - 1	SB-9-(18-20)	5173	0.9353	-	7471B	PPB	05 Dec 2018 15:27:20	U	SMPL
J6214-09 - 1	SB-4-(0-2)	29053	5.2992	-	7471B	PPB	05 Dec 2018 15:29:38	U	SMPL
J6214-11 - 1	SB-4-(3-5)	2665	0.4770	-	7471B	PPB	05 Dec 2018 15:31:55	U	SMPL
J6214-13 - 1	SB-1-(0-2)	31170	5.6861	-	7471B	PPB	05 Dec 2018 15:34:14	U	SMPL
J6214-15 - 1	SB-1-(10-12)	5218	0.9436	-	7471B	PPB	05 Dec 2018 15:36:32	U	SMPL
J6214-17 - 1	SB-10-(0-2)	20777	3.7868	-	7471B	PPB	05 Dec 2018 15:38:49	U	SMPL
CCV86 - 1	CCV86	26701	4.8694	-	7471B	PPB	05 Dec 2018 15:41:05	U	SMPL
CCB86 - 1	CCB86	-104	-0.0290	-	7471B	PPB	05 Dec 2018 15:43:22	U	SMPL
J6224-01 - 1	S-1	2253	0.4017	-	7471B	PPB	05 Dec 2018 15:45:39	U	SMPL
J6224-01DUP - 1	S-1DUP	2017	0.3586	-	7471B	PPB	05 Dec 2018 15:47:56	U	SMPL
J6224-01MS - 1	S-1MS	21365	3.8943	-	7471B	PPB	05 Dec 2018 15:50:12	U	SMPL
J6224-01MSD - 1	S-1MSD	21782	3.9705	-	7471B	PPB	05 Dec 2018 15:52:29	U	SMPL
J6224-01LX5 - 1		351	0.0541	-	7471B	PPB	05 Dec 2018 15:54:46	U	SMPL
J6224-01A - 1		24197	4.4118	-	7471B	PPB	05 Dec 2018 15:57:04	U	SMPL
CCV87 - 1	CCV87	26347	4.8047	-	7471B	PPB	05 Dec 2018 15:59:21	U	SMPL
CCB87 - 1	CCB87	-105	-0.0292	-	7471B	PPB	05 Dec 2018 16:01:39	U	SMPL