

LB99608


7470A

INSTRUMENT ID : CV1

Linear

 μ Abs. :

37679



A= 0.0000e+000

B= 2.6250e-004

C= -3.9480e-002

Rho= 0.9996867

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.042	-0.042	-8	0.000	-8				
0.2	0.200	0.164	-0.036	776	0.0 %	776				
2.5	2.500	2.499	-0.001	9669	0.0 %	9669				
5.0	5.000	5.104	0.104	19593	0.0 %	19593				
7.5	7.500	7.624	0.124	29193	0.0 %	29193				
10.0	10.000	9.851	-0.149	37679	0.0 %	37679				

LB99608 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Dec 2018 10:12:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-8	-	0.0000	7470A	PPB	06 Dec 2018 10:14:55	S	Std
0.2 - 1	50.2	776	-	0.2000	7470A	PPB	06 Dec 2018 10:17:04	S	Std
2.5 - 1	52.5	9669	-	2.5000	7470A	PPB	06 Dec 2018 10:19:12	S	Std
5.0 - 1	55	19593	-	5.0000	7470A	PPB	06 Dec 2018 10:21:21	S	Std
7.5 - 1	57.5	29193	-	7.5000	7470A	PPB	06 Dec 2018 10:23:29	S	Std
10.0 - 1	59.0	37679	-	10.0000	7470A	PPB	06 Dec 2018 10:28:04	S	Std
ICV08 - 1	ICV08	14776	3.8392	-	7470A	PPB	06 Dec 2018 10:30:41	U	SMPL
ICB08 - 1	ICB08	-93	-0.0639	-	7470A	PPB	06 Dec 2018 10:32:48	U	SMPL
CCV88 - 1	CCV88	19317	5.0313	-	7470A	PPB	06 Dec 2018 10:34:54	U	SMPL
CCB88 - 1	CCB88	-89	-0.0628	-	7470A	PPB	06 Dec 2018 10:37:01	U	SMPL
CRA - 1	CRA	764	0.1611	-	7470A	PPB	06 Dec 2018 10:39:08	U	SMPL
HighStd - 1	HighStd	37415	9.7820	-	7470A	PPB	06 Dec 2018 10:41:15	U	SMPL
ChkStd - 1	ChkStd	26839	7.0058	-	7470A	PPB	06 Dec 2018 10:43:22	U	SMPL
PB115381BL - 1	PBW	-143	-0.0770	-	7470A	PPB	06 Dec 2018 10:45:30	U	SMPL
PB115381BS - 1	LCSW	13752	3.5704	-	7470A	PPB	06 Dec 2018 10:47:39	U	SMPL
J6188-02 - 1	BU-03-120418-A	-84	-0.0615	-	7470A	PPB	06 Dec 2018 10:49:47	U	SMPL
J6188-02DUP - 1	BU-03-120418-ADUP	-31	-0.0476	-	7470A	PPB	06 Dec 2018 10:51:55	U	SMPL
J6188-02MS - 1	BU-03-120418-AMS	12490	3.2392	-	7470A	PPB	06 Dec 2018 10:54:03	U	SMPL
J6188-02MSD - 1	BU-03-120418-AMSD	12758	3.3095	-	7470A	PPB	06 Dec 2018 10:56:12	U	SMPL
J6194-02 - 1	OK-02-120418-A	-24	-0.0458	-	7470A	PPB	06 Dec 2018 10:58:19	U	SMPL
CCV89 - 1	CCV89	19602	5.1061	-	7470A	PPB	06 Dec 2018 11:00:25	U	SMPL
CCB89 - 1	CCB89	-86	-0.0621	-	7470A	PPB	06 Dec 2018 11:02:32	U	SMPL
J6194-04 - 1	OK-02-120418-B	-48	-0.0521	-	7470A	PPB	06 Dec 2018 11:04:39	U	SMPL
J6202-02 - 1	SAMPLE-1	-2	-0.0400	-	7470A	PPB	06 Dec 2018 11:06:46	U	SMPL
J6208-02 - 1	TP-1	630	0.1259	-	7470A	PPB	06 Dec 2018 11:08:53	U	SMPL
J6208-06 - 1	TP-2	476	0.0855	-	7470A	PPB	06 Dec 2018 11:11:01	U	SMPL
J6208-10 - 1	TP-3	73	-0.0203	-	7470A	PPB	06 Dec 2018 11:13:08	U	SMPL
J6208-14 - 1	TP-4	277	0.0332	-	7470A	PPB	06 Dec 2018 11:15:16	U	SMPL
J6208-18 - 1	TP-5	987	0.2196	-	7470A	PPB	06 Dec 2018 11:17:24	U	SMPL
J6208-22 - 1	TP-6	195	0.0117	-	7470A	PPB	06 Dec 2018 11:19:32	U	SMPL
J6210-08 - 1	C-1-112918	385	0.0616	-	7470A	PPB	06 Dec 2018 11:21:41	U	SMPL
J6210-10 - 1	C-2-112918	-13	-0.0429	-	7470A	PPB	06 Dec 2018 11:23:48	U	SMPL
CCV90 - 1	CCV90	19112	4.9774	-	7470A	PPB	06 Dec 2018 11:25:55	U	SMPL
CCB90 - 1	CCB90	-75	-0.0592	-	7470A	PPB	06 Dec 2018 11:28:02	U	SMPL
J6219-04 - 1	NM-342-135-120418	0	-0.0395	-	7470A	PPB	06 Dec 2018 11:30:09	U	SMPL
J6219-10 - 1	NM-342-135-120418-DEBRI	-5	-0.0408	-	7470A	PPB	06 Dec 2018 11:32:16	U	SMPL
J6233-02 - 1	SB-1	-9	-0.0418	-	7470A	PPB	06 Dec 2018 11:34:24	U	SMPL
J6233-04 - 1	SB-2	10	-0.0369	-	7470A	PPB	06 Dec 2018 11:36:31	U	SMPL
PB115381TB - 1	PB115329TB	31	-0.0313	-	7470A	PPB	06 Dec 2018 11:38:39	U	SMPL
J6188-02LX5 - 1		-20	-0.0447	-	7470A	PPB	06 Dec 2018 11:40:47	U	SMPL
J6188-02A - 1		14892	3.8697	-	7470A	PPB	06 Dec 2018 11:42:55	U	SMPL
CCV91 - 1	CCV91	19169	4.9924	-	7470A	PPB	06 Dec 2018 11:45:03	U	SMPL
CCB91 - 1	CCB91	-68	-0.0573	-	7470A	PPB	06 Dec 2018 11:47:12	U	SMPL