



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : jignesh

Run Number: LB99610

Slope : 98.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.1	4.01	12/05/2018	16:30
2	CAL2	1	Water			20.2	7.00	12/05/2018	16:32
3	CAL3	1	Water			20.2	10.04	12/05/2018	16:34
4	ICV	1	Water			20.2	7.01	12/05/2018	16:36
5	CCV1	1	Water			20.1	2.00	12/05/2018	17:00
6	J6243-03	1	Water			11.7	5.80	12/05/2018	17:02
7	J6243-03DUP	1	Water			11.8	5.81	12/05/2018	17:04
8	J6243-06	1	Water			12.4	5.50	12/05/2018	17:06
9	CCV2	1	Water			20.3	12.00	12/05/2018	17:08

WORKLIST(Hardcopy Internal Chain)

1399610

WorkList Name : pH120518

WorkList ID : 119964

Date : 12/5/2018 4:16:04 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/05/2018	Water	J6243-03	pH	Cool 4 deg C	PSEG01	L31	OILY-WATER-A	12/05/2018	9040C
12/05/2018	Water	J6243-06	pH	Cool 4 deg C	PSEG01	L31	OILY-WATER-B	12/05/2018	9040C

Date/Time 12/05/18 4:20
 Received by: SP
 Relinquished by: CP

Date/Time 12/05/18 5:15
 Received by: CP
 Relinquished by: PSp