

=====

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : ds

12/6/2018 12:16

Test: Total CN

Sample Id	<i>Mg/d</i> Result	Dil. 1 +	Response	Errors
ICV1	92.652	0.0	0.080	
ICB1	1.689	0.0	0.004	
CCV1	247.345	0.0	0.209	
CCB1	1.627	0.0	0.004	
PB115397BL	1.981	0.0	0.004	
PB115397BS	94.534	0.0	0.081	
MIDPB115397	250.103	0.0	0.211	
J6067-12	28.281	0.0	0.026	
J6067-12DUP	28.613	0.0	0.026	
J6086-07	20.024	0.0	0.019	
J6120-07	18.990	0.0	0.018	
J6144-07	20.796	0.0	0.020	
CCV2	260.039	0.0	0.219	
CCB2	1.690	0.0	0.004	

N	14
Mean	76.312
SD	100.0649
CV%	131.13

Aquakem v. 7.2AQ1

Results from time period:

Thu Dec 06 11:36:38 2018

Thu Dec 06 12:04:55 2018

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	1.4908	µg/l	12/6/2018 11:36:38	
5.0PPBCN	A	Total CN	P	5.9278	µg/l	12/6/2018 11:36:39	
10PPBCN	A	Total CN	P	11.3348	µg/l	12/6/2018 11:36:40	
50PPBCN	A	Total CN	P	50.7321	µg/l	12/6/2018 11:36:41	
100PPBCN	A	Total CN	P	96.5241	µg/l	12/6/2018 11:36:42	
250PPBCN	A	Total CN	P	246.8086	µg/l	12/6/2018 11:36:43	
500PPBCN	A	Total CN	P	502.1817	µg/l	12/6/2018 11:36:44	
ICV1	S	Total CN	P	92.6519	µg/l	12/6/2018 12:01:02	
ICB1	S	Total CN	P	1.6894	µg/l	12/6/2018 12:01:03	
CCV1	S	Total CN	P	247.3455	µg/l	12/6/2018 12:01:04	
CCB1	S	Total CN	P	1.627	µg/l	12/6/2018 12:01:05	
PB115397BL	S	Total CN	P	1.9813	µg/l	12/6/2018 12:01:06	
PB115397BS	S	Total CN	P	94.534	µg/l	12/6/2018 12:01:07	
MIDPB115397	S	Total CN	P	250.1034	µg/l	12/6/2018 12:01:08	
J6067-12	S	Total CN	P	28.2808	µg/l	12/6/2018 12:01:09	
J6067-12DUP	S	Total CN	P	28.6132	µg/l	12/6/2018 12:01:10	
J6086-07	S	Total CN	P	20.0245	µg/l	12/6/2018 12:01:11	
J6120-07	S	Total CN	P	18.9895	µg/l	12/6/2018 12:01:12	
J6144-07	S	Total CN	P	20.7962	µg/l	12/6/2018 12:04:53	
CCV2	S	Total CN	P	260.0388	µg/l	12/6/2018 12:04:54	
CCB2	S	Total CN	P	1.6898	µg/l	12/6/2018 12:04:55	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : als

12/5/2018 9:46


Test Total CN

Accepted 12/5/2018 9:46

Factor 1179
Bias 0.002

Coeff. of det. 0.999978

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	1.6396	0.0000	
2	5.0PPBCN	0.007	4.9581	5.0000	
3	10PPBCN	0.010	9.3408	10.0000	
4	50PPBCN	0.044	49.2158	50.0000	
5	100PPBCN	0.087	100.1577	100.0000	
6	250PPBCN	0.214	249.2549	250.0000	
7	500PPBCN	0.427	500.4330	500.0000	