

LB99616


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

57118



A= 0.0000e+000

B= 1.7504e-004

C= -2.3029e-002

Rho= 0.9999849

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	18	0.000	18				
0.2	0.200	0.187	-0.013	1199	0.0 %	1199				
2.5	2.500	2.528	0.028	14574	0.0 %	14574				
5.0	5.000	5.018	0.018	28799	0.0 %	28799				
7.5	7.500	7.512	0.012	43051	0.0 %	43051				
10.0	10.000	9.975	-0.025	57118	0.0 %	57118				

LB99616 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Dec 2018 14:02:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	18	-	0.0000	7471B	PPB	06 Dec 2018 14:04:44	S	Std
0.2 - 1	50.2	1199	-	0.2000	7471B	PPB	06 Dec 2018 14:07:03	S	Std
2.5 - 1	52.5	14574	-	2.5000	7471B	PPB	06 Dec 2018 14:09:21	S	Std
5.0 - 1	55	28799	-	5.0000	7471B	PPB	06 Dec 2018 14:11:39	S	Std
7.5 - 1	57.5	43051	-	7.5000	7471B	PPB	06 Dec 2018 14:13:57	S	Std
10.0 - 1	510	57118	-	10.0000	7471B	PPB	06 Dec 2018 14:16:15	S	Std
ICV09 - 1	ICV09	22015	3.8304	-7471B	PPB	06 Dec 2018 14:21:53	U	SMPL	
ICB09 - 1	ICB09	-85	-0.0379	-7471B	PPB	06 Dec 2018 14:24:09	U	SMPL	
CCV92 - 1	CCV92	27629	4.8131	-7471B	PPB	06 Dec 2018 14:26:25	U	SMPL	
CCB92 - 1	CCB92	-122	-0.0444	-7471B	PPB	06 Dec 2018 14:28:41	U	SMPL	
CRA - 1	CRA	1121	0.1732	-7471B	PPB	06 Dec 2018 14:30:58	U	SMPL	
HighStd - 1	HighStd	54934	9.5925	-7471B	PPB	06 Dec 2018 14:33:14	U	SMPL	
ChkStd - 1	ChkStd	39422	6.8773	-7471B	PPB	06 Dec 2018 14:35:30	U	SMPL	
PB115382BL - 1	PBS	-149	-0.0491	-7471B	PPB	06 Dec 2018 14:37:49	U	SMPL	
PB115382BS - 1	LCSS	18535	3.2213	-7471B	PPB	06 Dec 2018 14:40:07	U	SMPL	
J6188-01 - 1	BU-03-120418-A	2071	0.3395	-7471B	PPB	06 Dec 2018 14:42:24	U	SMPL	
J6188-01DUP - 1	BU-03-120418-ADUP	1952	0.3186	-7471B	PPB	06 Dec 2018 14:44:41	U	SMPL	
J6188-01MS - 1	BU-03-120418-AMS	21790	3.7910	-7471B	PPB	06 Dec 2018 14:46:59	U	SMPL	
J6188-01MSD - 1	BU-03-120418-AMSD	22038	3.8344	-7471B	PPB	06 Dec 2018 14:49:17	U	SMPL	
J6194-01 - 1	OK-02-120418-A	2187	0.3598	-7471B	PPB	06 Dec 2018 14:51:33	U	SMPL	
CCV93 - 1	CCV93	27807	4.8442	-7471B	PPB	06 Dec 2018 14:53:50	U	SMPL	
CCB93 - 1	CCB93	-141	-0.0477	-7471B	PPB	06 Dec 2018 14:56:05	U	SMPL	
J6194-03 - 1	OK-02-120418-B	801	0.1172	-7471B	PPB	06 Dec 2018 14:58:23	U	SMPL	
J6195-01 - 1	OR-01-120518-A	1552	0.2486	-7471B	PPB	06 Dec 2018 15:00:39	U	SMPL	
J6195-03 - 1	OR-01-120518-B	1846	0.3001	-7471B	PPB	06 Dec 2018 15:02:55	U	SMPL	
J6199-01 - 1	SU-01-120518-A	697	0.0990	-7471B	PPB	06 Dec 2018 15:05:11	U	SMPL	
J6199-03 - 1	SU-01-120518-B	507	0.0657	-7471B	PPB	06 Dec 2018 15:07:28	U	SMPL	
J6202-01 - 1	SAMPLE-1	1784	0.2892	-7471B	PPB	06 Dec 2018 15:09:45	U	SMPL	
J6214-18 - 1	SB-10-(13-15)	4000	0.6771	-7471B	PPB	06 Dec 2018 15:12:03	U	SMPL	
J6214-19 - 1	SB-9-(0-2)	27827	4.8477	-7471B	PPB	06 Dec 2018 15:14:20	U	SMPL	
J6214-21 - 1	SB-8-(0-2)	10360	1.7904	-7471B	PPB	06 Dec 2018 15:16:38	U	SMPL	
J6214-23 - 1	SB-3-(0-2)	33080	5.7672	-7471B	PPB	06 Dec 2018 15:18:55	U	SMPL	
CCV94 - 1	CCV94	27798	4.8426	-7471B	PPB	06 Dec 2018 15:21:12	U	SMPL	
CCB94 - 1	CCB94	-141	-0.0477	-7471B	PPB	06 Dec 2018 15:23:30	U	SMPL	
J6214-25 - 1	SB-3-(12.5-14.5)	12741	2.2071	-7471B	PPB	06 Dec 2018 15:25:47	U	SMPL	
J6214-27 - 1	SB-2-(0-2)	33686	5.8733	-7471B	PPB	06 Dec 2018 15:28:04	U	SMPL	
J6214-29 - 1	SB-2-(12.5-14.5)	53149	9.2800	-7471B	PPB	06 Dec 2018 15:30:20	U	SMPL	
J6219-01 - 1	NM-342-135-120418	-198	-0.0577	-7471B	PPB	06 Dec 2018 15:32:38	U	SMPL	
J6233-01 - 1	SB-1	688	0.0974	-7471B	PPB	06 Dec 2018 15:34:57	U	SMPL	
J6233-03 - 1	SB-2	140	0.0015	-7471B	PPB	06 Dec 2018 15:37:15	U	SMPL	
J6234-01 - 1	VNJ-218	35780	6.2398	-7471B	PPB	06 Dec 2018 15:39:32	U	SMPL	
J6237-01 - 1	SAMPLE-2	808	0.1184	-7471B	PPB	06 Dec 2018 15:41:50	U	SMPL	
J6242-01 - 1	PAINT-CHIPS	3385	0.5695	-7471B	PPB	06 Dec 2018 15:44:09	U	SMPL	
CCV95 - 1	CCV95	27771	4.8379	-7471B	PPB	06 Dec 2018 15:46:25	U	SMPL	
CCB95 - 1	CCB95	-107	-0.0418	-7471B	PPB	06 Dec 2018 15:48:42	U	SMPL	
J6188-01LX5 - 1		366	0.0410	-7471B	PPB	06 Dec 2018 15:51:00	U	SMPL	
J6188-01A - 1		24658	4.2930	-7471B	PPB	06 Dec 2018 15:53:16	U	SMPL	
CCV96 - 1	CCV96	27949	4.8691	-7471B	PPB	06 Dec 2018 15:55:33	U	SMPL	
CCB96 - 1	CCB96	-117	-0.0435	-7471B	PPB	06 Dec 2018 15:57:50	U	SMPL	