



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99622

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 051106

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 7                 | 27.0       | 80.6      | 80.6            | 12/06/2018 | 09:45     |
| 2   | J6243-04    |               | 1  | 9                 | >100.0     | >212.0    | >212.0          | 12/06/2018 | 10:20     |
| 3   | J6243-04DUP |               | 1  | 9                 | >100.0     | >212.0    | >212.0          | 12/06/2018 | 10:50     |
| 4   | J6243-07    |               | 1  | 7                 | >100.0     | >212.0    | >212.0          | 12/06/2018 | 11:20     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

LB99622

## WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp 12618

WorkList ID : 119996

Date : 12/6/2018 9:06:12 AM

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
| 12/10/2018 | Water  | J6243-04 | Flash Point | Cool 4 deg C | PSEG01   | L31              | OILY-WATER-A    | 12/05/2018   | 1010A  |
| 12/10/2018 | Water  | J6243-07 | Flash Point | Cool 4 deg C | PSEG01   | L31              | OILY-WATER-B    | 12/05/2018   | 1010A  |

Date/Time 12/6/18 9:15 AM  
Received by: KS  
Relinquished by: CP

Date/Time 12/6/18 11:25 AM  
Received by: CP  
Relinquished by: KS