



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB99623

Reagent/Standard	Lot/Log #
COD Digestion Vials Low Level 0-150Mg/L	W2422
COD ICV-LCS std, 50ppm	WP69885
COD calibration std. 100 ppm	WP69882
COD calibration std. 10 ppm	WP69881
COD calibration std. 150 ppm	WP69883
COD calibration std. 50 ppm	WP69884
COD calibration std. 0 ppm	WP69878
COD ICV-LCS std, 50ppm	WP69885
COD ICV-LCS std, 50ppm	WP70068
COD CCV std, 50ppm	WP70069

Temp In (C) : 150	Date In: 12/06/2018	Time In: 10:00
Temp Out (C) : 150	Date Out: 12/06/2018	Time Out: 12:00

Intercept: -0.7935

Slope: 1.0096

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	11/29/2018	12:45
2	CAL2	10	1	Water	9.000	11/29/2018	12:45
3	CAL3	50	1	Water	49.000	11/29/2018	12:46
4	CAL4	100	1	Water	100.000	11/29/2018	12:46
5	CAL5	150	1	Water	151.000	11/29/2018	12:47



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB99623

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	50.311	11/29/2018	12:47
2	ICB				1	Water	0.000	0.786	11/29/2018	12:48
3	CCV1	50			1	Water	50.000	50.311	12/06/2018	13:00
4	CCB1				1	Water	0.000	0.786	12/06/2018	13:00
5	LB99623BL				1	Water	0.000	0.786	12/06/2018	13:01
6	LB99623BS	50			1	Water	50.000	50.311	12/06/2018	13:01
7	J6130-01				1	Water	5.000	5.738	12/06/2018	13:02
8	J6130-01DUP				1	Water	5.000	5.738	12/06/2018	13:02
9	J6130-01MS	50			1	Water	54.000	54.272	12/06/2018	13:03
10	J6130-01MSD	50			1	Water	55.000	55.263	12/06/2018	13:03
11	J6130-02				1	Water	7.000	7.719	12/06/2018	13:04
12	J6207-01				1	Water	132.000	131.531	12/06/2018	13:04
13	J6215-01				1	Water	3.000	3.757	12/06/2018	13:05
14	J6241-02				1	Water	2.000	2.767	12/06/2018	13:05
15	CCV2	50			1	Water	51.000	51.301	12/06/2018	13:06
16	CCB2				1	Water	0.000	0.786	12/06/2018	13:06
17	J6243-03				1	Water	105.000	104.788	12/06/2018	13:07
18	J6243-06				10	Water	17.000	17.624	12/06/2018	13:07
19	CCV3	50			1	Water	51.000	51.301	12/06/2018	13:08
20	CCB3				1	Water	0.000	0.786	12/06/2018	13:08