

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/07/2018

TEMP1 IN: 103 12/06/2018 16:00 TEMP OUT: 104 12/07/2018 07:00

TEMP2 IN: 104 12/07/2018 07:40 TEMP OUT: 104 12/07/2018 09:00

TEMP3 IN: 104 12/07/2018 10:30 TEMP OUT: 104 12/07/2018 12:00

TEMP4 IN: 105 12/07/2018 12:40 TEMP OUT: 105 12/07/2018 14:20

Run Number: LB99637

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99637BL	LB99637BL	1.3865	1.3865	100	1.3865	1.3865	1.3865	0.0000	0
2	LB99637BS	LB99637BS	1.4010	1.4010	100	1.4505	1.4505	1.4505	0.0495	495
3	J6255-01	120518-JETT-F-D-B	1.4086	1.4086	200	1.4225	1.4225	1.4225	0.0139	69.5
4	J6255-01DUP	120518-JETT-F-D-BDUP	1.4158	1.4158	200	1.4299	1.4299	1.4299	0.0141	70.5
5	J6255-02	120518-JETT-F-D-S	1.3826	1.3826	250	1.3982	1.3982	1.3982	0.0156	62.4
6	J6255-03	120518-JETT-F-U-B	1.4064	1.4064	250	1.4212	1.4212	1.4212	0.0148	59.2
7	J6255-04	120518-JETT-F-U-S	1.4086	1.4086	250	1.4231	1.4231	1.4231	0.0145	58
8	J6255-05	120518-DBRI-F-D-B	1.4161	1.4161	250	1.4318	1.4318	1.4318	0.0157	62.8
9	J6255-06	120518-DBRI-F-D-S	1.3811	1.3811	250	1.3950	1.3950	1.3950	0.0139	55.6
10	J6257-01	COMP	1.4078	1.4078	500	1.4174	1.4174	1.4174	0.0096	19.2
11	J6269-03	OILY-WATER	1.4020	1.4020	100	1.4181	1.4181	1.4181	0.0161	161
12	J6286-01	15220032-1	1.3918	1.3918	500	1.3981	1.3981	1.3981	0.0063	12.6
13	J6287-01	32220008-3	1.3898	1.3898	60	1.4001	1.4001	1.4001	0.0103	171.7



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/07/2018

TEMP1 IN: 103 12/06/2018 16:00 TEMP OUT: 104 12/07/2018 07:00

TEMP2 IN: 104 12/07/2018 07:40 TEMP OUT: 104 12/07/2018 09:00

TEMP3 IN: 104 12/07/2018 10:30 TEMP OUT: 104 12/07/2018 12:00

TEMP4 IN: 105 12/07/2018 12:40 TEMP OUT: 105 12/07/2018 14:20

Run Number: LB99637

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS120718-1

WorkList ID : 120036

Date : 12/7/2018 8:35:45 AM

2B99637

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/12/2018	Water	J6255-01	TSS	Cool 4 deg C	HDR103	L33	120518-JETT-F-D-B	12/05/2018	SM2540 D
12/12/2018	Water	J6255-02	TSS	Cool 4 deg C	HDR103	L33	120518-JETT-F-D-S	12/05/2018	SM2540 D
12/12/2018	Water	J6255-03	TSS	Cool 4 deg C	HDR103	L33	120518-JETT-F-U-B	12/05/2018	SM2540 D
12/12/2018	Water	J6255-04	TSS	Cool 4 deg C	HDR103	L33	120518-JETT-F-U-S	12/05/2018	SM2540 D
12/12/2018	Water	J6255-05	TSS	Cool 4 deg C	HDR103	L33	120518-DBRI-F-D-B	12/05/2018	SM2540 D
12/12/2018	Water	J6255-06	TSS	Cool 4 deg C	HDR103	L33	120518-DBRI-F-D-S	12/05/2018	SM2540 D
12/12/2018	Water	J6257-01	TSS	Cool 4 deg C	ARAM01	L41	COMP	12/05/2018	SM2540 D
12/12/2018	Water	J6269-03	TSS	Cool 4 deg C	PSEG01	L31	OILY-WATER	12/06/2018	SM2540 D
12/13/2018	Water	J6286-01	TSS	Cool 4 deg C	PHAR02	L41	15220032-1	12/06/2018	SM2540 D
12/13/2018	Water	J6287-01	TSS	Cool 4 deg C	PHAR02	L41	32220008-3	12/06/2018	SM2540 D

Date/Time 12/6/18 09:00
 Received by: AS
 Relinquished by: CS

Date/Time 12/6/18 13:00
 Received by: AS
 Relinquished by: AS