



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : jignesh

Run Number: LB99640

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.5	4.01	12/07/2018	10:10
2	CAL2	1	Water			20.2	7.00	12/07/2018	10:12
3	CAL3	1	Water			20.3	10.04	12/07/2018	10:14
4	ICV	1	Water			20.4	7.01	12/07/2018	10:16
5	CCV1	1	Water			20.5	2.00	12/07/2018	10:18
6	J6260-03	1	Solid	20.02	20	21.3	8.13	12/07/2018	10:20
7	J6260-03DUP	1	Solid	20.03	20	21.4	8.14	12/07/2018	10:22
8	J6260-06	1	Solid	20.02	20	21.4	7.93	12/07/2018	10:24
9	J6260-09	1	Solid	20.04	20	21.3	7.50	12/07/2018	10:26
10	J6260-12	1	Solid	20.03	20	21.4	7.75	12/07/2018	10:28
11	J6261-01	1	Solid	20.04	20	22.1	5.94	12/07/2018	10:30
12	J6269-02	1	Solid	20.02	20	22.1	6.60	12/07/2018	10:32
13	CCV2	1	Water			20.5	12.00	12/07/2018	10:34

WORKLIST(Hardcopy Internal Chain)

JB99640

WorkList Name : corrosivity120618

WorkList ID : 120017

Date : 12/6/2018 5:06:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/04/2018	Solid	J6260-03	Corrosivity	Cool 4 deg C	LIRO01		SB-01-COMP	12/04/2018	9045D
12/04/2018	Solid	J6260-06	Corrosivity	Cool 4 deg C	LIRO01		SB-02-COMP	12/04/2018	9045D
12/04/2018	Solid	J6260-09	Corrosivity	Cool 4 deg C	LIRO01		SB-03-COMP	12/04/2018	9045D
12/04/2018	Solid	J6260-12	Corrosivity	Cool 4 deg C	LIRO01		SB-04-COMP	12/04/2018	9045D
12/06/2018	Solid	J6261-01	Corrosivity	Cool 4 deg C	PSEG01	L31	OILY-DEBRIS	12/06/2018	9045D
12/06/2018	Solid	J6269-02	Corrosivity	Cool 4 deg C	PSEG01	L31	OILY-SOIL	12/06/2018	9045D

Date/Time 12/07/18 10:00
 Received by: SP
 Relinquished by: CP

Date/Time 12/07/18 10:40
 Received by: CP
 Relinquished by: SP