



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99641

Ambient Barometric Pressure (mmHg): 765

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.3	12/07/2018	11:15
2	J6269-04		1	8	>100.0	>212.0	>212.0	12/07/2018	11:40
3	J6269-04DUP		1	8	>100.0	>212.0	>212.0	12/07/2018	12:02

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB99641

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp12718

WorkList ID : 120058

Date : 12/7/2018 10:45:06 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/12/2018	Water	J6269-04	C Flash Point	Cool 4 deg C	PSEG01	L31	OILY-WATER	12/06/2018	1010A

Date/Time 12/7/18 10:50 AM
Received by: KS
Relinquished by: CP

Date/Time 12/7/18 12:05 PM
Received by: CP
Relinquished by: KS