


245.1

LB99660
INSTRUMENT ID: CX1

Linear

 μ Abs.:

35030



A= 0.0000e+000

B= 2.8300e-004

C= -1.4936e-002

Rho= 0.9998419

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	-24	0.000	-24				
0.05	0.050	0.050	0.000	231	0.0 %	231				
0.2	0.200	0.215	0.015	811	0.0 %	811				
2.5	2.500	2.453	-0.047	8719	0.0 %	8719				
5.0	5.000	5.027	0.027	17816	0.0 %	17816				
7.5	7.500	7.629	0.129	27009	0.0 %	27009				
10.0	10.000	9.899	-0.101	35030	0.0 %	35030				

LB99660 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 07 Dec 2018 14:27:44

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-24	-	0.0000245.1	PPB	07 Dec 2018 14:29:53	S	Std
0.05 - 1	50.05	231	-	0.0500245.1	PPB	07 Dec 2018 14:32:11	S	Std
0.2 - 1	50.2	811	-	0.2000245.1	PPB	07 Dec 2018 14:42:54	S	Std
2.5 - 1	52.5	8719	-	2.5000245.1	PPB	07 Dec 2018 14:45:11	S	Std
5.0 - 1	55	17816	-	5.0000245.1	PPB	07 Dec 2018 14:47:29	S	Std
7.5 - 1	57.5	27009	-	7.5000245.1	PPB	07 Dec 2018 14:49:45	S	Std
10.0 - 1	510	35030	-	10.0000245.1	PPB	07 Dec 2018 15:00:04	S	Std
ICV11 - 1	ICV11	14108	3.9776	-245.1	PPB	07 Dec 2018 15:02:47	U	SMPL
ICB11 - 1	ICB11	-140	-0.0546	-245.1	PPB	07 Dec 2018 15:05:02	U	SMPL
CCV09 - 1	CCV09	18104	5.1085	-245.1	PPB	07 Dec 2018 15:07:19	U	SMPL
CCB09 - 1	CCB09	-141	-0.0548	-245.1	PPB	07 Dec 2018 15:09:34	U	SMPL
CRA - 1	CRA	197	0.0408	-245.1	PPB	07 Dec 2018 15:11:52	U	SMPL
HighStd - 1	HighStd	34594	9.7752	-245.1	PPB	07 Dec 2018 15:14:07	U	SMPL
ChkStd - 1	ChkStd	25520	7.2072	-245.1	PPB	07 Dec 2018 15:16:24	U	SMPL
J6225-01 - 1	120318-JETT-E-D-B	-256	-0.0874	-245.1	PPB	07 Dec 2018 15:18:42	U	SMPL
J6225-02 - 1	120318-JETT-E-D-S	-126	-0.0506	-245.1	PPB	07 Dec 2018 15:21:00	U	SMPL
J6225-03 - 1	120318-JETT-E-U-B	-93	-0.0413	-245.1	PPB	07 Dec 2018 15:23:17	U	SMPL
J6225-04 - 1	120318-JETT-E-U-S	-106	-0.0449	-245.1	PPB	07 Dec 2018 15:25:34	U	SMPL
J6225-05 - 1	120318-TIPC-E-D-M	-98	-0.0427	-245.1	PPB	07 Dec 2018 15:27:52	U	SMPL
J6225-06 - 1	120318-TIPC-E-U-M	-102	-0.0438	-245.1	PPB	07 Dec 2018 15:30:10	U	SMPL
J6225-06DUP - 1	120318-TIPC-E-U-MDUP	-115	-0.0475	-245.1	PPB	07 Dec 2018 15:32:25	U	SMPL
CCV10 - 1	CCV10	17883	5.0460	-245.1	PPB	07 Dec 2018 15:34:42	U	SMPL
CCB10 - 1	CCB10	-129	-0.0514	-245.1	PPB	07 Dec 2018 15:36:57	U	SMPL
J6225-06MS - 1	120318-TIPC-E-U-MMS	11856	3.3403	-245.1	PPB	07 Dec 2018 15:39:14	U	SMPL
J6225-06MSD - 1	120318-TIPC-E-U-MMSD	13576	3.8271	-245.1	PPB	07 Dec 2018 15:41:31	U	SMPL
J6254-01 - 1	120418-TIPC-F-D-M	-206	-0.0732	-245.1	PPB	07 Dec 2018 15:43:48	U	SMPL
J6254-02 - 1	120418-TIPC-F-U-M	-121	-0.0492	-245.1	PPB	07 Dec 2018 15:46:04	U	SMPL
J6255-01 - 1	120518-JETT-F-D-B	-121	-0.0492	-245.1	PPB	07 Dec 2018 15:48:21	U	SMPL
J6255-02 - 1	120518-JETT-F-D-S	-124	-0.0500	-245.1	PPB	07 Dec 2018 15:50:38	U	SMPL
J6255-03 - 1	120518-JETT-F-U-B	-112	-0.0466	-245.1	PPB	07 Dec 2018 15:52:55	U	SMPL
J6255-04 - 1	120518-JETT-F-U-S	-117	-0.0480	-245.1	PPB	07 Dec 2018 15:55:13	U	SMPL
J6255-05 - 1	120518-DBRI-F-D-B	-110	-0.0461	-245.1	PPB	07 Dec 2018 15:57:31	U	SMPL
J6255-06 - 1	120518-DBRI-F-D-S	-113	-0.0469	-245.1	PPB	07 Dec 2018 15:59:47	U	SMPL
CCV11 - 1	CCV11	17851	5.0369	-245.1	PPB	07 Dec 2018 16:02:04	U	SMPL
CCB11 - 1	CCB11	-149	-0.0571	-245.1	PPB	07 Dec 2018 16:04:20	U	SMPL
J6255-07 - 1	120518-TIPC-E-D-M	-118	-0.0483	-245.1	PPB	07 Dec 2018 16:06:38	U	SMPL
J6255-08 - 1	120518-TIPC-E-U-M	-115	-0.0475	-245.1	PPB	07 Dec 2018 16:08:54	U	SMPL
PB115413BL - 1	PBW	-14	-0.0189	-245.1	PPB	07 Dec 2018 16:13:02	U	SMPL
PB115413BS - 1	LCSW	14873	4.1941	-245.1	PPB	07 Dec 2018 16:15:18	U	SMPL
J6225-06LX5 - 1		-128	-0.0512	-245.1	PPB	07 Dec 2018 16:17:35	U	SMPL
J6225-06A - 1		15158	4.2748	-245.1	PPB	07 Dec 2018 16:19:52	U	SMPL
CCV12 - 1	CCV12	17927	5.0584	-245.1	PPB	07 Dec 2018 16:22:10	U	SMPL
CCB12 - 1	CCB12	-146	-0.0563	-245.1	PPB	07 Dec 2018 16:24:28	U	SMPL