

LB 99666


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37663



A= 0.0000e+000

B= 2.6615e-004

C= 2.8953e-003

Rho= 0.9994216

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	8	0.000	8				
0.2	0.200	0.179	-0.021	661	0.0 %	661				
2.5	2.500	2.438	-0.062	9149	0.0 %	9149				
5.0	5.000	5.236	0.236	19661	0.0 %	19661				
7.5	7.500	7.316	-0.184	27477	0.0 %	27477				
10.0	10.000	10.027	0.027	37663	0.0 %	37663				

LB99666 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Dec 2018 16:38:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	8	-	0.0000	7470A	PPB	07 Dec 2018 16:38:54	S	Std
0.2 - 1	50.2	661	-	0.2000	7470A	PPB	07 Dec 2018 16:41:03	S	Std
2.5 - 1	52.5	9149	-	2.5000	7470A	PPB	07 Dec 2018 16:43:12	S	Std
5.0 - 1	55	19661	-	5.0000	7470A	PPB	07 Dec 2018 16:45:20	S	Std
7.5 - 1	57.5	27477	-	7.5000	7470A	PPB	07 Dec 2018 16:47:29	S	Std
10.0 - 1	510	37663	-	10.0000	7470A	PPB	07 Dec 2018 16:49:36	S	Std
ICV12 - 1	ICV12	14701	3.9155	-	7470A	PPB	07 Dec 2018 16:52:10	U	SMPL
ICB12 - 1	ICB12	-82	-0.0189	-	7470A	PPB	07 Dec 2018 16:54:17	U	SMPL
CCV13 - 1	CCV13	19652	5.2332	-	7470A	PPB	07 Dec 2018 16:56:24	U	SMPL
CCB13 - 1	CCB13	-92	-0.0216	-	7470A	PPB	07 Dec 2018 16:58:30	U	SMPL
CRA - 1	CRA	645	0.1746	-	7470A	PPB	07 Dec 2018 17:00:37	U	SMPL
HighStd - 1	HighStd	37183	9.8991	-	7470A	PPB	07 Dec 2018 17:02:44	U	SMPL
ChkStd - 1	ChkStd	25748	6.8557	-	7470A	PPB	07 Dec 2018 17:04:51	U	SMPL
PB115462BL - 1	PBW	-158	-0.0392	-	7470A	PPB	07 Dec 2018 17:06:59	U	SMPL
PB115462BS - 1	LCSW	12975	3.4562	-	7470A	PPB	07 Dec 2018 17:09:08	U	SMPL
J6148-02 - 1	TP-MH-F	-87	-0.0203	-	7470A	PPB	07 Dec 2018 17:11:16	U	SMPL
J6148-02DUP - 1	TP-MH-FDUP	-7	0.0010	-	7470A	PPB	07 Dec 2018 17:13:24	U	SMPL
J6148-02MS - 1	TP-MH-FMS	13662	3.6390	-	7470A	PPB	07 Dec 2018 17:15:32	U	SMPL
J6148-02MSD - 1	TP-MH-FMSD	14022	3.7348	-	7470A	PPB	07 Dec 2018 17:17:42	U	SMPL
J6148-04 - 1	TP-MH-J	-73	-0.0165	-	7470A	PPB	07 Dec 2018 17:19:48	U	SMPL
CCV14 - 1	CCV14	19898	5.2987	-	7470A	PPB	07 Dec 2018 17:21:55	U	SMPL
CCB14 - 1	CCB14	-63	-0.0139	-	7470A	PPB	07 Dec 2018 17:24:02	U	SMPL
PB115462TB - 1	BL115230TB	-1	0.0026	-	7470A	PPB	07 Dec 2018 17:26:08	U	SMPL
J6148-02LX5 - 1		-20	-0.0024	-	7470A	PPB	07 Dec 2018 17:28:15	U	SMPL
J6148-02A - 1		14371	3.8277	-	7470A	PPB	07 Dec 2018 17:30:22	U	SMPL
CCV15 - 1	CCV15	19314	5.1433	-	7470A	PPB	07 Dec 2018 17:32:30	U	SMPL
CCB15 - 1	CCB15	-82	-0.0189	-	7470A	PPB	07 Dec 2018 17:34:37	U	SMPL