



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/10/2018

TEMP1 IN: 103 12/10/2018 07:00 TEMP OUT: 104 12/10/2018 08:05

TEMP2 IN: 104 12/10/2018 08:45 TEMP OUT: 104 12/10/2018 10:00

TEMP3 IN: 105 12/10/2018 11:00 TEMP OUT: 105 12/10/2018 12:35

TEMP4 IN: 105 12/10/2018 13:20 TEMP OUT: 105 12/10/2018 15:00

Run Number: LB99677

BalanceID: wc-sc-3

OvenID: wc-oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99677BL	LB99677BL	1.4038	1.4038	100	1.4038	1.4038	1.4038	0.0000	0
2	LB99677BS	LB99677BS	1.3956	1.3956	100	1.4460	1.4460	1.4460	0.0504	504
3	J6214-55	TW-2	1.3965	1.3965	300	1.4001	1.4001	1.4001	0.0036	12
4	J6307-01	001M-WILLETS-PT-BLVD (DEC)	1.3895	1.3895	100	1.3967	1.3967	1.3967	0.0072	72
5	J6307-02	002M-35TH-AVE (DEC)	1.3872	1.3872	200	1.3975	1.3975	1.3975	0.0103	51.5
6	J6308-01	001M-WILLETS-PT-BLVD (NOV)	1.4080	1.4080	200	1.4163	1.4163	1.4163	0.0083	41.5
7	J6308-02	002M-35TH-AVE (NOV)	1.3916	1.3916	200	1.3976	1.3976	1.3976	0.0060	30
8	J6312-01	EFF-WW	1.4063	1.4063	100	1.4139	1.4139	1.4139	0.0076	76
9	J6318-01	WEST-8TH-BOX-COVER-TEST-PIT	1.3932	1.3932	50	1.4095	1.4095	1.4095	0.0163	326
10	J6318-01DUP	WEST-8TH-BOX-COVER-TEST-PIT	1.4075	1.4075	50	1.4239	1.4239	1.4239	0.0164	328

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS121018

WorkList ID : 120095

Date : 12/10/2018 7:43:33 AM

AB99677

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/13/2018	Water	J6214-55	TSS	Cool 4 deg C	AKRF02		TW-2	12/06/2018	SM2540 D
12/11/2018	Water	J6307-01	TSS	1:1 HNO3 to p	TULL01	L11	001M-WILLETS-PT-BLVD(DE)	12/06/2018	SM2540 D
12/11/2018	Water	J6307-02	TSS	1:1 HNO3 to p	TULL01	L11	002M-35TH-AVE(DEC)	12/06/2018	SM2540 D
12/11/2018	Water	J6308-01	TSS	Cool 4 deg C	TULL01	L23	001M-WILLETS-PT-BLVD(NO)	12/06/2018	SM2540 D
12/11/2018	Water	J6308-02	TSS	Cool 4 deg C	TULL01	L23	002M-35TH-AVE(NOV)	12/06/2018	SM2540 D
12/14/2018	Water	J6312-01	TSS	Cool 4 deg C	ARDM01	L31	EFF-WW	12/07/2018	SM2540 D
12/14/2018	Water	J6318-01	TSS	Cool 4 deg C	TULL02	L22	WEST-8TH-BOX-COVER-TE	12/07/2018	SM2540 D

fp-12791048

Date/Time 12/10/18, 08:00
 Received by: AS
 Relinquished by: CP

Date/Time 12/10/18, 12:00
 Received by: [Signature]
 Relinquished by: AS