



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: AparnaDate: 12/11/2018OVEN TEMP1: 103 °C TEMP: 105 °CDATE & TIME IN: 12/10/2018 16:00 OUT: 12/11/2018 07:10OVEN TEMP2: 105 °C TEMP: 105 °CDATE & TIME IN: 12/11/2018 08:30 OUT: 12/11/2018 10:00Run Number: LB99683BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB99683BL	LB99683BL	82.6535	82.6535	100	82.6535	82.6535	82.6535	0.0000	0
2	J6214-55	TW-2	84.5827	84.5827	100	84.7353	84.7353	84.7353	0.1526	1526
3	J6318-01	WEST-8TH-BOX-COVER-TEST-PIT	85.0901	85.0901	100	85.2175	85.2175	85.2175	0.1274	1274
4	J6318-01DUP	WEST-8TH-BOX-COVER-TEST-PIT	74.7846	74.7846	100	74.9135	74.9135	74.9135	0.1289	1289

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ts121018

WorkList ID : 120107

Date : 12/10/2018 8:13:55 AM

12/10/18, 16:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/13/2018	Water	J6214-55	TS	Cool 4 deg C	AKRF02	L22	TW-2	12/06/2018	SM2540 B
12/14/2018	Water	J6318-01	TS	Cool 4 deg C	TULL02	L22	WEST-8TH-BOX-COVER-TE	12/07/2018	SM2540 B

Date/Time

Received by:

Relinquished by:

12/10/18, 14:00

12/10/18, 16:30